



DM-CAL-5E

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 C5E1122

 2022-10-12

 2022-12-10

Applicable Instruments	Parameter	Unit	Reference Value
DYMIND DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.33
	RBC	$\times 10^{12}/L$	4.52
	HGB	g/L	133
	MCV	fL	84.4
	PLT	$\times 10^9/L$	252

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.