

DM-3D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

2022-10-14

2023-02-10

| Applicable Instrument | Parameter | Unit | LOT B1122L | LOT B1122N | LOT B1122H |
|--|---------------------|----------------------|------------|-------------------|------------|
| DYMIND DH31 DH33 DH36 (Technical File Version A5.0 or higher) DH31 Vet DH36 Vet (Technical File Version A1.0 to A4.0) | WBC | ×10 ⁹ /L | / / | 8.10 ±1.00 | / / |
| | Lym% | % | / / | 32.2 ±8.0 | / / |
| | Gran% | % | / / | 60.5 ±8.0 | / / |
| | Mid% | % | / / | 7.3 ±6.0 | / / |
| | Lym# | ×10 ⁹ /L | / / | 2.61 ±0.70 | / / |
| | Gran# | ×10 ⁹ /L | / / | 4.90 ±0.70 | / / |
| | Mid# | ×10 ⁹ /L | / / | 0.59 ±0.50 | / / |
| | RBC | ×10 ¹² /L | / / | 4.82 ±0.24 | / / |
| | HGB | g/L | / / | 137 ±6 | / / |
| | HCT | % | / / | 42.9 ±3.0 | / / |
| | MCV | fL | / / | 89.0 ±5.0 | / / |
| | MCH | pg | / / | 28.4 ±2.5 | / / |
| | MCHC | g/L | / / | 320 ±30 | / / |
| | RDW-CV | % | / / | 17.2 ±3.0 | / / |
| | RDW-SD | fL | / / | 51.6 ±8.0 | / / |
| | PLT | ×10 ⁹ /L | / / | 262 ±40 | / / |
| | MPV | fL | / / | 10.1 ±3.0 | / / |
| | PDW | fL | / / | 13.2 ±3.0 | / / |
| | PCT | % | / / | 0.265 ±0.200 | / / |
| | P-LCR | % | / / | 28.6 ±10.0 | / / |
| P-LCC | ×10 ⁹ /L | / / | 75 ±25 | / / | |
| PDW | / | / / | 16.2 ±3.0 | / / | |

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B1122N

DM-3D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

📅 2022-10-14

🕒 2023-02-10

| Applicable Instrument | Parameter | Unit | LOT B1122L | LOT B1122N | LOT B1122H |
|---|---------------------|----------------------|------------|-------------------|------------|
| DYMIND DH20 DH21 DH22 DH23 DH25 DH26 (Technical File Version A1.0 or higher) | WBC | ×10 ⁹ /L | / / | 8.02 ±1.00 | / / |
| | Lym% | % | / / | 31.8 ±8.0 | / / |
| | Gran% | % | / / | 64.0 ±8.0 | / / |
| | Mid% | % | / / | 4.2 ±4.2 | / / |
| | Lym# | ×10 ⁹ /L | / / | 2.55 ±0.70 | / / |
| | Gran# | ×10 ⁹ /L | / / | 5.13 ±0.70 | / / |
| | Mid# | ×10 ⁹ /L | / / | 0.34 ±0.34 | / / |
| | RBC | ×10 ¹² /L | / / | 4.71 ±0.24 | / / |
| | HGB | g/L | / / | 135 ±6 | / / |
| | HCT | % | / / | 39.0 ±3.0 | / / |
| | MCV | fL | / / | 82.8 ±5.0 | / / |
| | MCH | pg | / / | 28.6 ±2.5 | / / |
| | MCHC | g/L | / / | 345 ±30 | / / |
| | RDW-CV | % | / / | 15.4 ±3.0 | / / |
| | RDW-SD | fL | / / | 47.7 ±8.0 | / / |
| | PLT | ×10 ⁹ /L | / / | 267 ±40 | / / |
| | MPV | fL | / / | 10.7 ±3.0 | / / |
| | PDW | fL | / / | 12.2 ±3.0 | / / |
| | PCT | % | / / | 0.285 ±0.200 | / / |
| | P-LCR | % | / / | 31.9 ±10.0 | / / |
| P-LCC | ×10 ⁹ /L | / / | 85 ±25 | / / | |

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B1122N

DM-3D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

2022-10-14

2023-02-10

| Applicable Instrument | Parameter | Unit | LOT B1122L | LOT B1122N | LOT B1122H |
|---|------------|----------------------|------------|-------------------|------------|
| DYMIND DH31 Vet DH36 Vet (Technical File Version A5.0 or higher) | WBC | ×10 ⁹ /L | / / | 7.73 ±1.00 | / / |
| | Lym% | % | / / | 30.4 ±8.0 | / / |
| | Gran% | % | / / | 62.6 ±8.0 | / / |
| | Mid% | % | / / | 7.0 ±6.0 | / / |
| | Lym# | ×10 ⁹ /L | / / | 2.35 ±0.70 | / / |
| | Gran# | ×10 ⁹ /L | / / | 4.84 ±0.70 | / / |
| | Mid# | ×10 ⁹ /L | / / | 0.54 ±0.50 | / / |
| | RBC | ×10 ¹² /L | / / | 4.56 ±0.24 | / / |
| | HGB | g/L | / / | 130 ±6 | / / |
| | HCT | % | / / | 41.5 ±3.0 | / / |
| | MCV | fL | / / | 91.0 ±5.0 | / / |
| | MCH | pg | / / | 28.5 ±2.5 | / / |
| | MCHC | g/L | / / | 313 ±30 | / / |
| | RDW-CV | % | / / | 14.8 ±3.0 | / / |
| | RDW-SD | fL | / / | 52.7 ±8.0 | / / |
| | PLT | ×10 ⁹ /L | / / | 255 ±40 | / / |
| | MPV | fL | / / | 11.9 ±3.0 | / / |
| | PDW | fL | / / | 12.3 ±3.0 | / / |
| | PCT | % | / / | 0.303 ±0.200 | / / |

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B1122N

DM-3D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL

📅 2022-10-14

🕒 2023-02-10

| Applicable Instrument | Parameter | Unit | LOT B1122L | LOT B1122N | LOT B1122H |
|--|---------------------|----------------------|------------|-------------------|------------|
| DYMIND DH36X (Technical File Version A1.0 or higher) | WBC | ×10 ⁹ /L | / / | 8.18 ±1.00 | / / |
| | Lym% | % | / / | 36.1 ±8.0 | / / |
| | Gran% | % | / / | 59.3 ±8.0 | / / |
| | Mid% | % | / / | 4.6 ±4.6 | / / |
| | Lym# | ×10 ⁹ /L | / / | 2.95 ±0.70 | / / |
| | Gran# | ×10 ⁹ /L | / / | 4.85 ±0.70 | / / |
| | Mid# | ×10 ⁹ /L | / / | 0.38 ±0.38 | / / |
| | RBC | ×10 ¹² /L | / / | 4.67 ±0.24 | / / |
| | HGB | g/L | / / | 135 ±6 | / / |
| | HCT | % | / / | 39.9 ±3.0 | / / |
| | MCV | fL | / / | 85.5 ±5.0 | / / |
| | MCH | pg | / / | 28.8 ±2.5 | / / |
| | MCHC | g/L | / / | 338 ±30 | / / |
| | RDW-CV | % | / / | 15.7 ±3.0 | / / |
| | RDW-SD | fL | / / | 50.1 ±8.0 | / / |
| | PLT | ×10 ⁹ /L | / / | 262 ±40 | / / |
| | MPV | fL | / / | 10.4 ±3.0 | / / |
| | PDW | fL | / / | 14.6 ±3.0 | / / |
| | PCT | % | / / | 0.273 ±0.200 | / / |
| | P-LCR | % | / / | 35.9 ±10.0 | / / |
| P-LCC | ×10 ⁹ /L | / / | 94 ±25 | / / | |

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B1122N