



# DM-CAL-5E

## HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

**CALIBRATOR**
 C5E0223

 2023-01-10

 2023-03-10

Applicable Instruments	Parameter	Unit	Reference Value
DYMIND  DH-600 DH-602 DH-605 DH-610 DH-612 DH-615  (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.11
	RBC	$\times 10^{12}/L$	4.59
	HGB	g/L	135
	MCV	fL	86.5
	PLT	$\times 10^9/L$	258

**【NOTE】**

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.