

	文件类型	技术文件	文件版本	V3.5.0
	项目代号	/	页 码	第 0 页 共 1 页
	文件编号	GL DM-56-24-en		

文件名称:

CALIBRATOR C5E0523 2023-06-10

适用范围: 1116 项目

编制:

审核:

批准:

日期:

日期:

日期:

修 订 记 录

版本	TCN/ECR/PCN 编号	修订内容概要	修改人	批准人	批准日期
V3.5.0		新建	黄江妮	郑剑通	2023.04.28

发至:

总经办 质量部 财务部 营销部 客服 计划 采购 工程 仓库

设备研发 试剂研发 物料质量 设备生产 试剂生产 试剂质检 设备质检

由试剂研发


DM-CAL-5E


HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

 C5E0523

 2023-04-12

 2023-06-10

Applicable Instruments	Parameter	Unit	Reference Value
DYMIND DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A1.6 or higher)	WBC	$\times 10^9/L$	9.50
	RBC	$\times 10^{12}/L$	4.44
	HGB	g/L	133
	MCV	fL	88.5
	PLT	$\times 10^9/L$	265

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.