

	文件类型	技术文件	文件版本	V9.6.0	
	项目代号	/	页 码	第 0 页 共 1 页	
	文件编号	GL DM-56-05-en			
文件名称： <p style="text-align: center;">CALIBRATOR N0623 2023-07-10</p> 适用范围：公司所有三分类仪器					
编制：		审核：		批准：	
日期：		日期：		日期：	
修 订 记 录					
版本	TCN/ECR/PCN 编号	修订内容概要	修改人	批准人	批准日期
V9.6.0	/	新建	王雨婷	郑剑通	2023.05.26
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由试剂研发					

DM-CAL

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
LOT N0623

 2023-05-12

 2023-07-10

Applicable Instruments	Parameter	Unit	Reference Value
DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31 Vet, DH36 Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	9.21
	RBC	$\times 10^{12}/L$	4.72
	HGB	g/L	126
	MCV	fL	88.6
	PLT	$\times 10^9/L$	255
DH20, DH21, DH22 DH23, DH25, DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.09
	RBC	$\times 10^{12}/L$	4.60
	HGB	g/L	127
	MCV	fL	82.9
	PLT	$\times 10^9/L$	251
DH31 Vet, DH36 Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	9.31
	RBC	$\times 10^{12}/L$	4.58
	HGB	g/L	126
	MCV	fL	89.9
	PLT	$\times 10^9/L$	255
DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.53
	RBC	$\times 10^{12}/L$	4.73
	HGB	g/L	131
	MCV	fL	85.4
	PLT	$\times 10^9/L$	249

【NOTE】

1. The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.