
	文件类型	技术文件	文件版本	V9.4.0		
	项目代号	/	页 码	第 0 页 共 1 页		
	文件编号	GL DM-56-01- en				
文件名称： <p style="text-align: center;"><b>CONTROL DM-5D BC0723 2023-09-10 for 7003</b></p> 适用范围： 7003 项目						
编制：		审核：		批准：		
日期：		日期：		日期：		
修 订 记 录						
版本	TCN/ECR/PCN 编号	修订内容概要		修改人	批准人	批准日期
V9.4.0	/	新建		段利超	郑剑通	2023.07.06
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库					
	<input checked="" type="checkbox"/> 设备研发 <input type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input checked="" type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input type="checkbox"/> 试剂质检 <input type="checkbox"/> 设备质检					
由试剂研发						


# DM-5D

## HEMATOLOGY CONTROL

Reference Values provided by DYMIND

**CONTROL**

 2023-06-13

 2023-09-10

Applicable Instrument	Parameter	Unit	LOT BC0723L	LOT BC0723N	LOT BC0723H
<b>DYMIND</b>  DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.1 or higher)	<b>WBC</b>	×10 <sup>9</sup> /L	<b>3.45</b> ±0.50	<b>8.06</b> ±1.00	<b>18.82</b> ±2.50
	Neu%	%	50.2 ±9.0	55.5 ±8.0	63.2 ±7.0
	Lym%	%	39.8 ±9.0	32.1 ±8.0	21.3 ±6.0
	Mon%	%	3.2 ±3.2	3.3 ±3.3	3.9 ±3.9
	Eos%	%	6.8 ±5.0	9.1 ±6.0	11.6 ±7.0
	Bas%	%	6.5 ±6.5	4.2 ±4.2	1.6 ±1.6
	Neu#	×10 <sup>9</sup> /L	1.73 ±0.40	4.47 ±0.70	11.89 ±1.40
	Lym#	×10 <sup>9</sup> /L	1.37 ±0.40	2.59 ±0.70	4.01 ±1.10
	Mon#	×10 <sup>9</sup> /L	0.11 ±0.11	0.27 ±0.27	0.73 ±0.73
	Eos#	×10 <sup>9</sup> /L	0.23 ±0.15	0.73 ±0.50	2.18 ±1.30
	Bas#	×10 <sup>9</sup> /L	0.22 ±0.22	0.34 ±0.34	0.30 ±0.30
	<b>RBC</b>	×10 <sup>12</sup> /L	<b>2.37</b> ±0.18	<b>4.67</b> ±0.24	<b>5.33</b> ±0.50
	<b>HGB</b>	g/L	<b>56</b> ±4	<b>132</b> ±6	<b>173</b> ±8
	HCT	%	17.6 ±2.0	39.2 ±3.0	49.2 ±4.0
	<b>MCV</b>	fL	<b>74.5</b> ±5.0	<b>83.9</b> ±5.0	<b>92.3</b> ±6.0
	MCH	pg	23.7 ±2.5	28.1 ±2.5	32.3 ±2.5
	MCHC	g/L	316 ±30	334 ±30	348 ±30
	RDW-CV	%	15.3 ±3.0	13.2 ±3.0	12.1 ±3.0
	RDW-SD	fL	52.2 ±10.0	50.6 ±10.0	51.2 ±12.0
	<b>PLT</b>	×10 <sup>9</sup> /L	<b>66</b> ±20	<b>272</b> ±40	<b>499</b> ±60
MPV	fL	8.9 ±3.0	8.8 ±3.0	9.1 ±3.0	
PDW	fL	11.0 ±3.0	12.1 ±3.0	12.6 ±3.0	
PCT	%	0.059 ±0.050	0.239 ±0.100	0.454 ±0.200	

**【NOTE】**

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



BC0723L



BC0723N



BC0723H