
	文件类型	技术文件	文件版本	V4.4.0	
	项目代号	DM3113	页 码	第 0 页 共 3 页	
	文件编号	GL DM-56-09-en			
文件名称: <p style="text-align: center;">Table of CRP Calibrator Reference Value</p>					
适用范围: <p style="text-align: center;">DM3113 项目</p>					
编制: 曾婷婷 日期: 2023.08.08		审核: 魏保立 日期: 2023.08.08		批准: 魏保立 日期: 2023.08.08	
修 订 记 录					
版本	TCN/ECR/P CN 编号	修订内容概要	修改人	批准人	批准日期
V4.4.0	/	新建	曾婷婷	魏保立	2023.08.08
发至:	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input type="checkbox"/> 营销部 <input type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input type="checkbox"/> 设备研发 <input type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input type="checkbox"/> 试剂质检 <input type="checkbox"/> 设备质 检				
由试剂研发					

Table of CRP Calibrator Reference Value

C-Reactive Protein Calibrator

Code: DM3113

 2023080101

 20240731

Model	Level (multi-point)	Reference Value (mg/L)
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.9)	Calibrator 1	4.6
	Calibrator 2	17.8
	Calibrator 3	50.4
	Calibrator 4	98.5
	Calibrator 5	138.9
	Calibrator 6	174.3
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.10 and above with heating)	Calibrator 1	5.2
	Calibrator 2	20.2
	Calibrator 3	48.3
	Calibrator 4	103.8
	Calibrator 5	150.6
	Calibrator 6	194.9
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.10 and above without heating)	Calibrator 1	6.1
	Calibrator 2	21.2
	Calibrator 3	48.8
	Calibrator 4	100.9
	Calibrator 5	153.7
	Calibrator 6	193.2
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A5.11 and below)	Calibrator 1	4.7
	Calibrator 2	21.6
	Calibrator 3	56.5
	Calibrator 4	109.6
	Calibrator 5	152.6
	Calibrator 6	206.3
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A6.0 and above with heating)	Calibrator 1	5.5
	Calibrator 2	22.9
	Calibrator 3	55.5
	Calibrator 4	105.8
	Calibrator 5	153.2
	Calibrator 6	195.0

DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A6.0 and above without heating)	Calibrator 1	5.6
	Calibrator 2	19.8
	Calibrator 3	50.9
	Calibrator 4	125.0
	Calibrator 5	176.8
	Calibrator 6	236.0
D6-CRP, DH76CRP (Applicable to technical document version B2.0)	Calibrator 1	5.2
	Calibrator 2	20.7
	Calibrator 3	54.3
	Calibrator 4	107.5
	Calibrator 5	156.4
	Calibrator 6	187.7
D6-CRP, DH76CRP (Applicable to technical document version B2.2 and above)	Calibrator 1	5.1
	Calibrator 2	20.7
	Calibrator 3	48.9
	Calibrator 4	104.1
	Calibrator 5	143.3
	Calibrator 6	178.6
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Applicable to technical document versions A7.5 and below with heating)	Calibrator 1	5.8
	Calibrator 2	21.2
	Calibrator 3	53.5
	Calibrator 4	108.6
	Calibrator 5	159.8
	Calibrator 6	196.7
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Applicable to technical document versions A7.5 and below without heating)	Calibrator 1	5.6
	Calibrator 2	21.0
	Calibrator 3	54.6
	Calibrator 4	112.5
	Calibrator 5	158.3
	Calibrator 6	206.5
DM79X, DM78X, DM72X (Applicable to technical document version A1.0 and above)	Calibrator 1	5.5
	Calibrator 2	21.6
	Calibrator 3	53.7
	Calibrator 4	109.2
	Calibrator 5	150.2
	Calibrator 6	196.8

DH36X、DH33X、DH31X (Applicable to technical document versions A1.0 and above)	Calibrator 1	5.2
	Calibrator 2	20.4
	Calibrator 3	53.4
	Calibrator 4	106.9
	Calibrator 5	150.2
	Calibrator 6	192.9