	文件类型	技术文件	文件版本	V10.1.0	
	项目代号	/	页 码	第 0 页 共 2 页	
	文件编号	GL DM-56-05-en			
文件名称： <p style="text-align: center;"><b>CALIBRATOR N0124 2024-02-10</b></p> 适用范围：公司所有三分类仪器					
编制：刘嘉 日期：2023.12.29		审核：刘金盈 日期：2023.12.29		批准：郑剑通 日期：2023.12.29	
修 订 记 录					
版本	TCN/ECR/PCN 编号	修订内容概要	修改人	批准人	批准日期
V10.1.0	/	新建	刘嘉	郑剑通	2023.12.29
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由试剂研发					

# DM-CAL

## HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

**CALIBRATOR**

LOT N0124

2023-12-13

2024-02-10

Applicable Instruments	Parameter	Unit	Reference Value
DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31 Vet, DH36 Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	9.29
	RBC	$\times 10^{12}/L$	4.67
	HGB	g/L	130
	MCV	fL	90.7
	PLT	$\times 10^9/L$	247
DH21, DH23 DH25, DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.17
	RBC	$\times 10^{12}/L$	4.54
	HGB	g/L	131
	MCV	fL	84.1
	PLT	$\times 10^9/L$	247
DH20, DH22 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.29
	RBC	$\times 10^{12}/L$	4.52
	HGB	g/L	132
	MCV	fL	87.5
	PLT	$\times 10^9/L$	244
DH31 Vet, DH36 Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	9.25
	RBC	$\times 10^{12}/L$	4.53
	HGB	g/L	127
	MCV	fL	91.3
	PLT	$\times 10^9/L$	242

**【NOTE】**

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.


# DM-CAL


## HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

**CALIBRATOR**

LOT N0124

 2023-12-13

 2024-02-10

Applicable Instruments	Parameter	Unit	Reference Value
DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.72
	RBC	$\times 10^{12}/L$	4.55
	HGB	g/L	132
	MCV	fL	87.8
	PLT	$\times 10^9/L$	245

**【NOTE】**

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.