
	文件类型	技术文件	文件版本	V4.7.0	
	项目代号	DM3113	页 码	第 0 页 共 2 页	
	文件编号	GL DM-56-08-en			
文件名称: <p style="text-align: center;"><b>Table of CRP Control Reference Value</b></p>					
适用范围: <p style="text-align: center;">DM3113 项目</p>					
编制: 黄子苒 日期: 2023.12.22		审核: 曾婷婷 日期: 2023.12.22		批准: 魏保立 日期: 2023.12.22	
修 订 记 录					
版本	TCN/ECR/P CN 编号	修订内容概要	修改人	批准人	批准日期
V4.7.0	/	新建	黄子苒	魏保立	2023.12.22
发至:	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input type="checkbox"/> 营销部 <input type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input type="checkbox"/> 设备研发 <input type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input type="checkbox"/> 试剂质检 <input type="checkbox"/> 设备质 检				
由试剂研发					

## Table of CRP Control Reference Value

### C-Reactive Protein Control

**Code: DM3113**

LOT 2023121301

 20241212

Model	Level	Reference Value (mg/L)	Reference Range (mg/L)	Deviation Range (mg/L)
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.9)	<b>Control 1</b>	2.8	2.0-3.6	±0.8
	<b>Control 2</b>	12.8	9.0-16.6	±3.8
	<b>Control 3</b>	31.4	22-40.8	±9.4
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.10 and above with heating)	<b>Control 1</b>	3.4	2.4-4.4	±1.0
	<b>Control 2</b>	14.2	9.9-18.5	±4.3
	<b>Control 3</b>	37.1	26.0-48.2	±11.1
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.10 and above without heating)	<b>Control 1</b>	3.4	2.4-4.4	±1.0
	<b>Control 2</b>	15.1	10.6-19.6	±4.5
	<b>Control 3</b>	38.6	27.0-50.2	±11.6
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A5.11 and below)	<b>Control 1</b>	3.1	2.2-4.0	±0.9
	<b>Control 2</b>	13.3	9.3-17.3	±4.0
	<b>Control 3</b>	33.1	23.2-43.0	±9.9
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A6.0 and above with heating)	<b>Control 1</b>	3.1	2.2-4.0	±0.9
	<b>Control 2</b>	14.0	9.8-18.2	±4.2
	<b>Control 3</b>	38.8	27.2-50.4	±11.6
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A6.0 and above without heating)	<b>Control 1</b>	3.5	2.4-4.6	±1.1
	<b>Control 2</b>	15.1	10.6-19.6	±4.5
	<b>Control 3</b>	37.6	26.3-48.9	±11.3

D6-CRP, DH76CRP (Applicable to technical document version B2.0)	<b>Control 1</b>	3.4	2.4-4.4	$\pm 1.0$
	<b>Control 2</b>	15.0	10.5-19.5	$\pm 4.5$
	<b>Control 3</b>	44.1	30.9-57.3	$\pm 13.2$
D6-CRP, DH76CRP (Applicable to technical document version B2.2 and above)	<b>Control 1</b>	3.0	2.1-3.9	$\pm 0.9$
	<b>Control 2</b>	15.3	10.7-19.9	$\pm 4.6$
	<b>Control 3</b>	39.5	27.6-51.4	$\pm 11.9$
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Applicable to technical document versions A7.6 and below with heating)	<b>Control 1</b>	3.3	2.3-4.3	$\pm 1.0$
	<b>Control 2</b>	14.1	9.9-18.3	$\pm 4.2$
	<b>Control 3</b>	38.9	27.2-50.6	$\pm 11.7$
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Applicable to technical document versions A7.6 and below without heating)	<b>Control 1</b>	3.5	2.4-4.6	$\pm 1.1$
	<b>Control 2</b>	14.2	9.9-18.5	$\pm 4.3$
	<b>Control 3</b>	38.0	26.6-49.4	$\pm 11.4$
DM79X, DM78X, DM72X (Applicable to technical document version A1.0 and above)	<b>Control 1</b>	4.1	2.9-5.3	$\pm 1.2$
	<b>Control 2</b>	15.5	10.8-20.2	$\pm 4.7$
	<b>Control 3</b>	39.7	27.8-51.6	$\pm 11.9$
DH36X, DH33X, DH31X (Applicable to technical document versions A1.0 and above)	<b>Control 1</b>	3.3	2.3-4.3	$\pm 1.0$
	<b>Control 2</b>	14.8	10.4-19.2	$\pm 4.4$
	<b>Control 3</b>	41.7	29.2-54.2	$\pm 12.5$
DP-H10, DP-H12, DP-H13, DP-H15, DP-H16, DP-H17 (Applicable to technical document versions A1.0 and above)	<b>Control 1</b>	3.4	2.4-4.4	$\pm 1.0$
	<b>Control 2</b>	16.3	11.4-21.2	$\pm 4.9$
	<b>Control 3</b>	38.1	26.7-49.5	$\pm 11.4$