
	文件类型	技术文件	文件版本	V4.7.0	
	项目代号	DM3113	页 码	第 0 页 共 3 页	
	文件编号	GL DM-56-09-en			
文件名称: <p style="text-align: center;">Table of CRP Calibrator Reference Value</p>					
适用范围: <p style="text-align: center;">DM3113 项目</p>					
编制: 黄子苾 日期: 2023.12.22		审核: 曾婷婷 日期: 2023.12.22		批准: 魏保立 日期: 2023.12.22	
修 订 记 录					
版本	TCN/ECR/P CN 编号	修订内容概要	修改人	批准人	批准日期
V4.7.0	/	新建	黄子苾	魏保立	2023.12.22
发至:	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input type="checkbox"/> 营销部 <input type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input type="checkbox"/> 设备研发 <input type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input type="checkbox"/> 试剂质检 <input type="checkbox"/> 设备质 检				
由试剂研发					

Table of CRP Calibrator Reference Value

C-Reactive Protein Calibrator

Code: DM3113

 2023121301

 20241212

Model	Level (multi-point)	Reference Value (mg/L)
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.9)	Calibrator 1	4.5
	Calibrator 2	16.8
	Calibrator 3	45.4
	Calibrator 4	108.2
	Calibrator 5	129.6
	Calibrator 6	177.5
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.10 and above with heating)	Calibrator 1	5.3
	Calibrator 2	19.3
	Calibrator 3	49.7
	Calibrator 4	99.6
	Calibrator 5	141.8
	Calibrator 6	190.1
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.10 and above without heating)	Calibrator 1	4.7
	Calibrator 2	19.6
	Calibrator 3	47.4
	Calibrator 4	109.2
	Calibrator 5	153.9
	Calibrator 6	203.0
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A5.11 and below)	Calibrator 1	4.6
	Calibrator 2	16.6
	Calibrator 3	44.2
	Calibrator 4	98.7
	Calibrator 5	142.8
	Calibrator 6	197.4
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A6.0 and above with heating)	Calibrator 1	4.7
	Calibrator 2	19.4
	Calibrator 3	51.3
	Calibrator 4	103.2
	Calibrator 5	154.3
	Calibrator 6	188.3

DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A6.0 and above without heating)	Calibrator 1	5.2
	Calibrator 2	19.5
	Calibrator 3	48.9
	Calibrator 4	107.5
	Calibrator 5	155.0
	Calibrator 6	193.6
D6-CRP, DH76CRP (Applicable to technical document version B2.0)	Calibrator 1	5.1
	Calibrator 2	19.2
	Calibrator 3	53.7
	Calibrator 4	112.0
	Calibrator 5	158.3
	Calibrator 6	206.0
D6-CRP, DH76CRP (Applicable to technical document version B2.2 and above)	Calibrator 1	5.4
	Calibrator 2	20.3
	Calibrator 3	50.7
	Calibrator 4	99.8
	Calibrator 5	146.9
	Calibrator 6	206.1
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Applicable to technical document versions A7.6 and below with heating)	Calibrator 1	5.0
	Calibrator 2	19.2
	Calibrator 3	52.3
	Calibrator 4	105.1
	Calibrator 5	152.2
	Calibrator 6	192.6
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Applicable to technical document versions A7.6 and below without heating)	Calibrator 1	5.1
	Calibrator 2	19.1
	Calibrator 3	47.8
	Calibrator 4	106.4
	Calibrator 5	150.7
	Calibrator 6	193.6
DM79X, DM78X, DM72X (Applicable to technical document version A1.0 and above)	Calibrator 1	5.5
	Calibrator 2	19.3
	Calibrator 3	53.2
	Calibrator 4	101.5
	Calibrator 5	145.1
	Calibrator 6	208.0

DH36X、DH33X、DH31X (Applicable to technical document versions A1.0 and above)	Calibrator 1	5.0
	Calibrator 2	19.5
	Calibrator 3	52.7
	Calibrator 4	113.7
	Calibrator 5	162.5
	Calibrator 6	214.6
DP-H10、DP-H12、DP-H13、DP-H15、DP-H16、 DP-H17 (Applicable to technical document versions A1.0 and above)	Calibrator 1	/
	Calibrator 2	/
	Calibrator 3	/
	Calibrator 4	/
	Calibrator 5	/
	Calibrator 6	/