

	文件类型	技术文件	文件版本	V10.1.0	
	项目代号	/	页 码	第 0 页 共 1 页	
	文件编号	GL DM-56-03-en			
文件名称： <p style="text-align: center;"><b>CONTROL DM-3D B0224 2024-05-10 for DP-H10 Vet</b></p> 适用范围： P102 A1.0 及以上					
编制：王雨婷 日期：2024.01.26		审核：刘金盈 日期：2024.01.26		批准：郑剑通 日期：2024.01.26	
修 订 记 录					
版本	TCN/ECR/PC N 编号	修订内容概要	修改人	批准人	批准日期
V10.1.0	/	新建	王雨婷	郑剑通	2024.01.26
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由试剂研发					

# DM-3D

## HEMATOLOGY CONTROL

Reference Values provided by DYMIND

**CONTROL**

📅 2024-01-12

🕒 2024-05-10

Applicable Instruments	Parameter	Unit	LOT B0224L	LOT B0224N	LOT B0224H
DYMIND  DP-H10 Vet (Technical File Version A1.0 or higher)	<b>WBC</b>	×10 <sup>9</sup> /L	<b>1.92</b> ±0.80	<b>7.77</b> ±2.00	<b>20.90</b> ±4.50
	Lym%	%	60.4 ±10.0	31.3 ±8.0	13.8 ±8.0
	Gran%	%	29.9 ±12.0	63.4 ±8.0	84.3 ±8.0
	Mid%	%	9.7 ±9.7	5.3 ±5.0	1.9 ±1.9
	Lym#	×10 <sup>9</sup> /L	1.16 ±0.60	2.43 ±1.00	2.88 ±2.00
	Gran#	×10 <sup>9</sup> /L	0.57 ±0.50	4.93 ±1.80	17.62 ±5.00
	Mid#	×10 <sup>9</sup> /L	0.19 ±0.19	0.41 ±0.41	0.40 ±0.40
	<b>RBC</b>	×10 <sup>12</sup> /L	<b>2.36</b> ±0.50	<b>4.53</b> ±1.20	<b>5.73</b> ±1.20
	<b>HGB</b>	g/L	<b>55</b> ±15	<b>127</b> ±30	<b>177</b> ±40
	HCT	%	18.9 ±5.0	42.0 ±10.0	57.4 ±15.0
	<b>MCV</b>	fL	<b>80.2</b> ±12.0	<b>92.7</b> ±8.0	<b>100.1</b> ±8.0
	MCH	pg	23.3 ±3.0	28.1 ±3.0	31.0 ±3.0
	MCHC	g/L	299 ±45	309 ±45	317 ±45
	RDW-CV	%	16.8 ±3.0	13.8 ±3.0	13.3 ±3.0
	RDW-SD	fL	52.5 ±8.0	49.8 ±8.0	51.6 ±8.0
	<b>PLT</b>	×10 <sup>9</sup> /L	<b>71</b> ±40	<b>276</b> ±80	<b>571</b> ±120
	MPV	fL	9.9 ±4.0	10.6 ±3.0	10.4 ±3.0
	PDW	fL	9.4 ±9.0	10.7 ±3.0	10.5 ±3.0
PCT	%	0.070 ±0.070	0.288 ±0.288	0.589 ±0.400	

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



B0224L



B0224N



B0224H