

	文件类型	技术文件	文件版本	V10.2.0		
	项目代号	/	页 码	第 0 页 共 1 页		
	文件编号	GL DM-56-05-en				
文件名称： <p style="text-align: center;">CALIBRATOR N0224 2024-03-10 for DP-H10 Vet</p> 适用范围： P102 A1.0 及以上						
编制：王雨婷 日期：2024.01.26		审核：刘金盈 日期：2024.01.26		批准：郑剑通 日期：2024.01.26		
修 订 记 录						
版本	TCN/ECR/PCN 编号	修订内容概要		修改人	批准人	批准日期
V10.2.0	/	新建		王雨婷	郑剑通	2024.01.26
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由试剂研发						

DM-CAL

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT N0224

2024-01-11

2024-03-10

Applicable Instruments	Parameter	Unit	Reference Value
DP-H10 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.11
	RBC	$\times 10^{12}/L$	4.58
	HGB	g/L	129
	MCV	fL	91.3
	PLT	$\times 10^9/L$	236

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.