





# DM-CAL

## HEMATOLOGY CALIBRATOR

LOT N0224

 2024-01-11 2024-03-10

Applicable Instruments	Parameter	Unit	Reference Value
<b>DYMIND</b> <b>DP-H10、</b> <b>DP-H12、</b> <b>DP-H13、</b> <b>DP-H15、</b> <b>DP-H16、</b> <b>DP-H17</b>  (Technical File Version A1.10 or higher) ;	<b>WBC</b>	$\times 10^9/L$	9.41
	<b>RBC</b>	$\times 10^{12}/L$	4.98
	<b>HGB</b>	g/L	139
	<b>MCV</b>	fL	87.7
	<b>PLT</b>	$\times 10^9/L$	275

**【NOTE】**

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.