

	文件类型	技术文件	文件版本	V5.1.0	
	项目代号	DM3113	页 码	第 0 页 共 3 页	
	文件编号	GL DM-56-09-en			
文件名称: <p style="text-align: center;">Table of CRP Calibrator Reference Value</p>					
适用范围: <p style="text-align: center;">DM3113 项目</p>					
编制: 曾婷婷 日期: 2024.01.29		审核: 田文飞 日期: 2024.01.29		批准: 田文飞 日期: 2024.01.29	
修 订 记 录					
版本	TCN/ECR/P CN 编号	修订内容概要	修改人	批准人	批准日期
V5.1.0	/	新建	黄子苒	田文飞	2024.01.29
发至:	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input type="checkbox"/> 营销部 <input type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input type="checkbox"/> 设备研发 <input type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input type="checkbox"/> 试剂质检 <input type="checkbox"/> 设备质检				
由试剂研发					

Table of CRP Calibrator Reference Value

C-Reactive Protein Calibrator

Code: DM3113

LOT 2024012201

 20250121

Model	Level (multi-point)	Reference Value (mg/L)
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.9)	Calibrator 1	5.1
	Calibrator 2	19.4
	Calibrator 3	46.2
	Calibrator 4	97.3
	Calibrator 5	126.8
	Calibrator 6	157.1
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.10 and above with heating)	Calibrator 1	5.5
	Calibrator 2	21.5
	Calibrator 3	46.3
	Calibrator 4	98.7
	Calibrator 5	138.6
	Calibrator 6	174.9
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.10 and above without heating)	Calibrator 1	5.5
	Calibrator 2	21.4
	Calibrator 3	46.4
	Calibrator 4	98.8
	Calibrator 5	134.6
	Calibrator 6	175.9
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A5.11 and below)	Calibrator 1	5.1
	Calibrator 2	22.8
	Calibrator 3	45.8
	Calibrator 4	91.5
	Calibrator 5	134.4
	Calibrator 6	166.3
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A6.0 and above with heating)	Calibrator 1	5.6
	Calibrator 2	21.6
	Calibrator 3	49.7
	Calibrator 4	109.1
	Calibrator 5	143.2
	Calibrator 6	168.1

DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A6.0 and above without heating)	Calibrator 1	5.9
	Calibrator 2	22.5
	Calibrator 3	51.1
	Calibrator 4	100.8
	Calibrator 5	134.6
	Calibrator 6	172.8
D6-CRP, DH76CRP (Applicable to technical document version B2.0)	Calibrator 1	5.5
	Calibrator 2	20.6
	Calibrator 3	51.2
	Calibrator 4	109.4
	Calibrator 5	137.0
	Calibrator 6	174.2
D6-CRP, DH76CRP (Applicable to technical document version B2.2 and above)	Calibrator 1	5.9
	Calibrator 2	22.1
	Calibrator 3	48.0
	Calibrator 4	100.4
	Calibrator 5	136.7
	Calibrator 6	171.8
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Applicable to technical document versions A8.0 and below with heating)	Calibrator 1	5.7
	Calibrator 2	22.6
	Calibrator 3	49.9
	Calibrator 4	110.3
	Calibrator 5	146.3
	Calibrator 6	175.9
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Applicable to technical document versions A8.0 and below without heating)	Calibrator 1	5.5
	Calibrator 2	22.6
	Calibrator 3	49.3
	Calibrator 4	110.6
	Calibrator 5	140.1
	Calibrator 6	171.4
DM79X, DM78X, DM72X (Applicable to technical document version A1.0 and above)	Calibrator 1	5.9
	Calibrator 2	22.0
	Calibrator 3	54.8
	Calibrator 4	105.2
	Calibrator 5	143.1
	Calibrator 6	170.2

DH36X、DH33X、DH31X (Applicable to technical document versions A1.0 and above)	Calibrator 1	5.8
	Calibrator 2	22.9
	Calibrator 3	50.0
	Calibrator 4	105.9
	Calibrator 5	138.7
	Calibrator 6	177.4
DP-H10、DP-H12、DP-H13、DP-H15、DP-H16、 DP-H17 (Applicable to technical document versions A1.0 and above)	Calibrator 1	5.4
	Calibrator 2	20.8
	Calibrator 3	44.2
	Calibrator 4	95.2
	Calibrator 5	135.3
	Calibrator 6	169.1