

	文件类型	技术文件	文件版本	V4.1.0	
	项目代号	/	页 码	第 0 页 共 1 页	
	文件编号	GL DM-56-18-en			
文件名称： <p style="text-align: center;">CONTROL DM-6D 6D0224 2024-04-10</p>					
适用范围：1116 项目					
编制：刘嘉 日期：2024.01.31		审核：刘金盈 日期：2024.01.31		批准：郑剑通 日期：2024.01.31	
修 订 记 录					
版本	TCN/ECR/PCN 编号	修订内容概要	修改人	批准人	批准日期
V4.1.0	/	新建	刘嘉	郑剑通	2024.01.31
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由试剂研发					

DM-6D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND


CONTROL
 2024-01-12

 2024-04-10


Instruments	Parameter	Unit	LOT	6D0224L	LOT	6D0224N	LOT	6D0224H
DYMIND DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	WBC	×10 ⁹ /L	4.34	±1.20	7.81	±1.80	22.23	±3.00
	Neu#	×10 ⁹ /L	3.07	±0.40	4.86	±0.80	16.00	±2.00
	Lym#	×10 ⁹ /L	0.62	±0.40	1.97	±0.70	4.06	±1.80
	Mon#	×10 ⁹ /L	0.37	±0.40	0.57	±0.50	0.90	±0.80
	Eos#	×10 ⁹ /L	0.28	±0.40	0.41	±0.70	1.27	±1.60
	Bas#	×10 ⁹ /L	3.44	±0.40	5.74	±0.70	16.82	±1.60
	IG#	×10 ⁹ /L	0.58	±0.50	0.99	±1.00	2.98	±2.50
	Neu%	%	70.8	±12.0	62.2	±12.0	72.0	±12.0
	Lym%	%	14.3	±10.0	25.2	±10.0	18.3	±10.0
	Mon%	%	8.5	±8.0	7.3	±8.0	4.0	±8.0
	Eos%	%	6.4	±10.0	5.3	±10.0	5.7	±10.0
	Bas%	%	79.2	±10.0	73.5	±10.0	75.7	±10.0
	IG%	%	13.4	±10.0	12.7	±10.0	13.4	±10.0
	RBC	×10 ¹² /L	2.25	±0.20	4.49	±0.30	5.50	±0.50
	HGB	g/L	61	±4	134	±6	176	±8
	HCT	%	17.9	±2.0	39.8	±2.5	52.2	±3.0
	MCV	fL	79.5	±5.0	88.6	±5.0	95.0	±5.0
	MCH	pg	27.1	±2.5	29.9	±2.5	32.0	±2.5
	MCHC	g/L	341	±30	337	±30	337	±30
	RDW-CV	%	15.9	±5.0	14.1	±5.0	14.1	±6.0
RDW-SD	fL	46.9	±10.0	46.5	±10.0	49.7	±12.0	
PLT	×10 ⁹ /L	77	±20	284	±45	560	±65	
MPV	fL	7.0	±3.0	7.1	±3.0	7.1	±3.0	
PDW	fL	7.1	±5.0	7.8	±5.0	8.0	±5.0	
PCT	%	0.054	±0.050	0.201	±0.100	0.397	±0.200	
P-LCR	%	10.7	±20.0	10.3	±20.0	10.3	±20.0	
P-LCC	×10 ⁹ /L	8	±10	29	±30	58	±50	
IPF	%	2.0	±2.0	1.8	±1.8	2.0	±2.0	

【NOTE】


1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



6D0224L



6D0224N



6D0224H