

	文件类型	技术文件	文件版本	V4.2.0
	项目代号	/	页 码	第 0 页 共 1 页
	文件编号	GL DM-56-24-en		

文件名称:

### CALIBRATOR C5E0224 2024-03-10

适用范围: 1116 项目

编制: 刘嘉

审核: 刘金盈

批准: 郑剑通

日期: 2024.01.31

日期: 2024.01.31

日期: 2024.01.31

#### 修 订 记 录

版本	TCN/ECR/PCN 编号	修订内容概要	修改人	批准人	批准日期
V4.2.0		新建	刘嘉	郑剑通	2024.01.31

发至:

总经办  质量部  财务部  营销部  客服  计划  采购  工程  仓库

设备研发  试剂研发  物料质量  设备生产  试剂生产  试剂质检  设备质检


由试剂研发


# DM-CAL-5E

## HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

**CALIBRATOR**
 C5E0224

 2024-01-11

 2024-03-10

Applicable Instruments	Parameter	Unit	Reference Value
DYMIND  DH-600 DH-602 DH-605 DH-610 DH-612 DH-615  (Technical File Version A2.0 or higher)	WBC	$\times 10^9/L$	9.29
	RBC	$\times 10^{12}/L$	4.45
	HGB	g/L	133
	MCV	fL	88.2
	PLT	$\times 10^9/L$	250

**【NOTE】**

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.