

	文件类型	技术文件	文件版本	V10.3.0		
	项目代号	/	页码	第0页 共1页		
	文件编号	GL DM-56-05-en				
文件名称： <p style="text-align: center;"><b>CALIBRATOR N0324 2024-04-10 for P102</b></p> 适用范围：P102项目						
编制：石丽岐 日期：2024.03.07		审核：刘金盈 日期：2024.03.07		批准：郑剑通 日期：2024.03.07		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V10.3.0	/	新建		石丽岐	郑剑通	2024.03.07
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input checked="" type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input type="checkbox"/> 试剂质检 <input type="checkbox"/> 设备质检					
由试剂研发						

# DM-CAL

## HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

**CALIBRATOR**

LOT N0324

2024-02-11

2024-04-10

Applicable Instrument	Parameter	Unit	Reference Value
DP-H10 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.29
	RBC	$\times 10^{12}/L$	4.55
	HGB	g/L	139
	MCV	fL	92.4
	PLT	$\times 10^9/L$	235

**【NOTE】**

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.