

	文件类型	技术文件	文件版本	V4.3.0
	项目代号	/	页 码	第 0 页 共 1 页
	文件编号	GL DM-56-24-en		

文件名称:

CALIBRATOR C5E0324 2024-04-10

适用范围: 1116 项目

编制: 石丽岐

审核: 刘金盈

批准: 郑剑通

日期: 2024.03.04

日期: 2024.03.04

日期: 2024.03.04

修 订 记 录

版本	TCN/ECR/PCN 编号	修订内容概要	修改人	批准人	批准日期
V4.3.0		新建	石丽岐	郑剑通	2024.03.04

发至:

总经办 质量部 财务部 营销部 客服 计划 采购 工程 仓库

设备研发 试剂研发 物料质量 设备生产 试剂生产 试剂质检 设备质检

由试剂研发

DM-CAL-5E

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT C5E0324

M 2024-02-11

E 2024-04-10

Applicable Instruments	Parameter	Unit	Reference Value
DYMIND DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	WBC	$\times 10^9/L$	9.45
	RBC	$\times 10^{12}/L$	4.45
	HGB	g/L	133
	MCV	fL	87.8
	PLT	$\times 10^9/L$	246

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.