	文件类型	技术文件	文件版本	V4.2.0
	项目代号	/	页 码	第 0 页 共 1 页
	文件编号	GL DM-56-18-en		

文件名称:

CONTROL DM-6D 6D0424 2024-06-10

适用范围: 1116 项目

编制: 石丽岐

审核: 刘金盈

批准: 郑剑通

日期: 2024.03.22

日期: 2024.03.22

日期: 2024.03.22

修 订 记 录

版本	TCN/ECR/PCN 编号	修订内容概要	修改人	批准人	批准日期
V4.2.0	/	新建	石丽岐	郑剑通	2024.03.22

发至:

总经办 质量部 财务部 营销部 客服 计划 采购 工程 仓库

设备研发 试剂研发 物料质量 设备生产 试剂生产 试剂质检 设备质检

由试剂研发

DM-6D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND

CONTROL


📅 2024-03-13

🕒 2024-06-10


Instruments	Parameter	Unit	LOT	6D0424L	LOT	6D0424N	LOT	6D0424H
DYMIND DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	WBC	×10 ⁹ /L	4.28	±1.20	7.66	±1.80	22.30	±3.00
	Neu#	×10 ⁹ /L	3.03	±0.40	4.95	±0.80	15.91	±2.00
	Lym#	×10 ⁹ /L	0.52	±0.40	1.83	±0.70	4.68	±1.80
	Mon#	×10 ⁹ /L	0.49	±0.40	0.65	±0.50	1.06	±0.80
	Eos#	×10 ⁹ /L	0.23	±0.40	0.23	±0.70	0.65	±1.60
	Bas#	×10 ⁹ /L	3.49	±0.40	5.67	±0.70	16.75	±1.60
	IG#	×10 ⁹ /L	0.57	±0.50	0.97	±1.00	2.97	±2.50
	Neu%	%	70.8	±12.0	64.6	±12.0	71.3	±12.0
	Lym%	%	12.2	±10.0	23.9	±10.0	21.0	±10.0
	Mon%	%	11.5	±8.0	8.4	±8.0	4.8	±8.0
	Eos%	%	5.4	±10.0	3.0	±10.0	2.9	±10.0
	Bas%	%	81.5	±10.0	74.0	±10.0	75.1	±10.0
	IG%	%	13.3	±10.0	12.6	±10.0	13.3	±10.0
	RBC	×10 ¹² /L	2.31	±0.20	4.48	±0.30	5.63	±0.50
	HGB	g/L	63	±4	131	±6	176	±8
	HCT	%	18.4	±2.0	38.7	±2.5	52.0	±3.0
	MCV	fL	79.6	±5.0	86.5	±5.0	92.3	±5.0
	MCH	pg	27.3	±2.5	29.3	±2.5	31.3	±2.5
	MCHC	g/L	343	±30	339	±30	339	±30
	RDW-CV	%	16.1	±5.0	14.9	±5.0	14.4	±6.0
	RDW-SD	fL	47.1	±10.0	47.4	±10.0	48.7	±12.0
	PLT	×10 ⁹ /L	72	±20	276	±45	531	±65
	MPV	fL	7.4	±3.0	7.2	±3.0	7.1	±3.0
	PDW	fL	9.1	±5.0	8.8	±5.0	8.9	±5.0
	PCT	%	0.053	±0.050	0.198	±0.100	0.378	±0.200
	P-LCR	%	12.0	±20.0	10.1	±20.0	9.5	±20.0
	P-LCC	×10 ⁹ /L	9	±10	28	±30	51	±50
IPF	%	2.3	±2.3	2.0	±2.0	1.6	±1.6	

【NOTE】


1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



6D0424L



6D0424N



6D0424H