

	文件类型	技术文件	文件版本	V2.2.0	
	项目代号	/	页 码	第 0 页 共 1 页	
	文件编号	GL DM-56-18-en			
文件名称： <p style="text-align: center;">CONTROL DM-6D 6D0424 2024-06-10</p> 适用范围：1117 项目					
编制：刘嘉 日期：2024.03.28		审核：刘金盈 日期：2024.03.28		批准：郑剑通 日期：2024.03.28	
修 订 记 录					
版本	TCN/ECR/PCN 编号	修订内容概要	修改人	批准人	批准日期
V2.2.0	/	新建	刘嘉	郑剑通	2024.03.28
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由试剂研发					

DM-6D

HEMATOLOGY CONTROL


Reference Values provided by DYMIND
CONTROL
 2024-03-13

 2024-06-10

Applicable Instruments	Parameter	Unit	LOT	6D0424L	LOT	6D0424N	LOT	6D0424H
DYMIND DH-800 series DH-800 CRP series DH-800 CS series (Technical File Version A1.1 or higher)	WBC	×10 ⁹ /L	3.91	±1.20	7.56	±1.80	21.29	±3.00
	Neu#	×10 ⁹ /L	2.76	±0.40	4.76	±0.80	15.11	±2.00
	Lym#	×10 ⁹ /L	0.39	±0.40	1.81	±0.70	4.15	±1.80
	Mon#	×10 ⁹ /L	0.43	±0.40	0.58	±0.50	0.88	±0.80
	Eos#	×10 ⁹ /L	0.24	±0.40	0.23	±0.70	0.64	±1.60
	Bas#	×10 ⁹ /L	0.09	±0.40	0.18	±0.70	0.51	±1.60
	IG#	×10 ⁹ /L	0.54	±0.50	0.96	±1.00	2.94	±2.50
	Neu%	%	70.6	±12.0	63.0	±12.0	71.0	±12.0
	Lym%	%	10.0	±10.0	23.9	±10.0	19.5	±10.0
	Mon%	%	11.0	±8.0	7.6	±8.0	4.1	±8.0
	Eos%	%	6.0	±10.0	3.0	±10.0	3.0	±10.0
	Bas%	%	2.3	±10.0	2.4	±10.0	2.4	±10.0
	IG%	%	13.8	±10.0	12.7	±10.0	13.8	±10.0
	RBC	×10 ¹² /L	2.34	±0.20	4.50	±0.30	5.66	±0.50
	HGB	g/L	63	±4	131	±6	176	±8
	HCT	%	19.8	±2.0	41.2	±2.5	55.0	±3.0
	MCV	fL	84.8	±5.0	91.7	±5.0	97.3	±5.0
	MCH	pg	26.8	±2.5	29.2	±2.5	31.1	±2.5
	MCHC	g/L	317	±30	318	±30	319	±30
	RDW-CV	%	17.2	±5.0	15.2	±5.0	14.4	±6.0
	RDW-SD	fL	50.8	±10.0	49.5	±10.0	50.2	±12.0
	PLT	×10 ⁹ /L	75	±20	260	±45	496	±65
	MPV	fL	8.3	±3.0	7.8	±3.0	7.8	±3.0
	PDW	fL	9.0	±5.0	8.4	±5.0	8.3	±5.0
	PCT	%	0.062	±0.050	0.204	±0.100	0.385	±0.200
	P-LCR	%	14.3	±20.0	11.6	±20.0	11.1	±20.0
	P-LCC	×10 ⁹ /L	11	±10	30	±30	55	±50
	IPF	%	12.6	±12.6	12.5	±12.5	12.4	±12.4
	NRBC#	×10 ⁹ /L	0.23	±0.40	0.20	±0.40	1.19	±1.50
	NRBC%	%	5.7	±6.0	2.0	±4.0	5.6	±6.0

【NOTE】

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



6D0424L



6D0424N



6D0424H