

	文件类型	技术文件	文件版本	V2.4.0
	项目代号	/	页 码	第 0 页 共 1 页
	文件编号	GL DM-56-61		

文件名称:

CALIBRATOR CPF0424 2024-05-10

适用范围: 1117 项目

编制: 刘嘉

审核: 刘金盈

批准: 郑剑通

日期: 2024.03.28

日期: 2024.03.28

日期: 2024.03.28

修 订 记 录

版本	TCN/ECR/PCN 编号	修订内容概要	修改人	批准人	批准日期
V2.4.0	/	新建	刘嘉	郑剑通	2024.03.28

发至:

总经办 质量部 财务部 营销部 客服 计划 采购 工程 仓库

设备研发 试剂研发 物料质量 设备生产 试剂生产 试剂质检 设备质检

由试剂研发

DM-CAL-PF

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT CPF0424

2024-03-12

2024-05-10

Applicable Instruments	Parameter	Unit	Reference Value
DYMIND DH-800 series DH-800 CRP series DH-800 CS series (Technical File Version A1.1 or higher)	PLT-F	$\times 10^9/L$	189

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.