	文件类型	技术文件	文件版本	V10.4.0		
	项目代号	/	页码	第0页 共2页		
	文件编号	GL DM-56-05-en				
文件名称： <p style="text-align: center;"><b>CALIBRATOR N0424 2024-05-10</b></p> 适用范围： 公司所有三分类仪器						
编制：石哲臣 日期：2024.03.28		审核：刘金盈 日期：2024.03.28		批准：郑剑通 日期：2024.03.28		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V10.4.0	/	新建		石哲臣	郑剑通	2024.03.28
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由试剂研发						


# DM-CAL


## HEMATOLOGY CALIBRATOR

**Reference Values provided by DYMIND**

<b>CALIBRATOR</b>
-------------------

LOT	N0424
-----	-------

	2024-03-12
---	------------

	2024-05-10
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31 Vet, DH36 Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	9.27
	RBC	$\times 10^{12}/L$	4.71
	HGB	g/L	135
	MCV	fL	93.3
	PLT	$\times 10^9/L$	238
DH21, DH23 DH25, DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.17
	RBC	$\times 10^{12}/L$	4.52
	HGB	g/L	135
	MCV	fL	86.4
	PLT	$\times 10^9/L$	239
DH20, DH22 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.35
	RBC	$\times 10^{12}/L$	4.58
	HGB	g/L	136
	MCV	fL	90.8
	PLT	$\times 10^9/L$	237
DH31 Vet, DH36 Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	9.20
	RBC	$\times 10^{12}/L$	4.52
	HGB	g/L	130
	MCV	fL	93.6
	PLT	$\times 10^9/L$	236

**【NOTE】**

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

# DM-CAL

## HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

**CALIBRATOR**

LOT N0424

2024-03-12

2024-05-10

Applicable Instruments	Parameter	Unit	Reference Value
DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.63
	RBC	$\times 10^{12}/L$	4.63
	HGB	g/L	136
	MCV	fL	89.9
	PLT	$\times 10^9/L$	234
DP-H10 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.19
	RBC	$\times 10^{12}/L$	4.52
	HGB	g/L	140
	MCV	fL	93.6
	PLT	$\times 10^9/L$	236

**【NOTE】**

- The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
- Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
- Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.