

	文件类型	技术文件	文件版本	V5.3.0	
	项目代号	DM3113	页 码	第 0 页 共 3 页	
	文件编号	GL DM-56-09-en			
文件名称: <p style="text-align: center;">Table of CRP Calibrator Reference Value</p>					
适用范围: <p style="text-align: center;">DM3113 项目</p>					
编制: 秦俊飞 日期: 2024.04.12		审核: 曾婷婷 日期: 2024.04.12		批准: 田文飞 日期: 2024.04.12	
修 订 记 录					
版本	TCN/ECR/P CN 编号	修订内容概要	修改人	批准人	批准日期
V5.3.0	/	新建	秦俊飞	田文飞	2024.04.12
发至:	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input type="checkbox"/> 营销部 <input type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input type="checkbox"/> 设备研发 <input type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input type="checkbox"/> 试剂质检 <input type="checkbox"/> 设备质检				
由试剂研发					

Table of CRP Calibrator Reference Value

C-Reactive Protein Calibrator

Code: DM3113
LOT 2024032101

 20250320

Model	Level (multi-point)	Reference Value (mg/L)
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.9)	Calibrator 1	5.1
	Calibrator 2	19.4
	Calibrator 3	47.4
	Calibrator 4	93.6
	Calibrator 5	131.9
	Calibrator 6	175.7
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.10 and above with heating)	Calibrator 1	5.3
	Calibrator 2	19.5
	Calibrator 3	50.6
	Calibrator 4	106.8
	Calibrator 5	154.8
	Calibrator 6	193.5
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.10 and above without heating)	Calibrator 1	6.3
	Calibrator 2	21.5
	Calibrator 3	52.5
	Calibrator 4	105.3
	Calibrator 5	155.6
	Calibrator 6	175.9
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A5.11 and below)	Calibrator 1	4.5
	Calibrator 2	19.8
	Calibrator 3	50.6
	Calibrator 4	106.5
	Calibrator 5	142.5
	Calibrator 6	175.3
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A6.0 and above with heating)	Calibrator 1	5.3
	Calibrator 2	21.5
	Calibrator 3	51.9
	Calibrator 4	109.2
	Calibrator 5	142.2
	Calibrator 6	185.5

DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A6.0 and above without heating)	Calibrator 1	6.3
	Calibrator 2	20.1
	Calibrator 3	52.6
	Calibrator 4	105.8
	Calibrator 5	131.0
	Calibrator 6	172.8
D6-CRP, DH76CRP (Applicable to technical document version B2.0)	Calibrator 1	5.0
	Calibrator 2	19.2
	Calibrator 3	49.2
	Calibrator 4	103.5
	Calibrator 5	131.8
	Calibrator 6	175.9
D6-CRP, DH76CRP (Applicable to technical document version B2.2 and above)	Calibrator 1	5.5
	Calibrator 2	19.9
	Calibrator 3	55.5
	Calibrator 4	104.3
	Calibrator 5	142.3
	Calibrator 6	180.6
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Applicable to technical document versions A8.1 and below with heating)	Calibrator 1	5.6
	Calibrator 2	20.8
	Calibrator 3	51.3
	Calibrator 4	109.3
	Calibrator 5	152.6
	Calibrator 6	178.7
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Applicable to technical document versions A8.1 and below without heating)	Calibrator 1	5.4
	Calibrator 2	20.1
	Calibrator 3	51.6
	Calibrator 4	105.3
	Calibrator 5	150.9
	Calibrator 6	191.3
DM79X, DM78X, DM72X (Applicable to technical document version A1.0 and above)	Calibrator 1	5.9
	Calibrator 2	20.5
	Calibrator 3	55.8
	Calibrator 4	105.6
	Calibrator 5	137.4
	Calibrator 6	181.0

DH36X、DH33X、DH31X (Applicable to technical document versions A1.0 and above)	Calibrator 1	5.8
	Calibrator 2	20.7
	Calibrator 3	55.1
	Calibrator 4	108.5
	Calibrator 5	147.6
	Calibrator 6	176.0
DP-H10、DP-H12、DP-H13、DP-H15、DP-H16、 DP-H17 (Applicable to technical document versions A1.0 and above)	Calibrator 1	5.3
	Calibrator 2	18.1
	Calibrator 3	48.7
	Calibrator 4	95.0
	Calibrator 5	140.2
	Calibrator 6	180.8