

| | | | | |
|---|------|----------------|------|-------------|
|  | 文件类型 | 技术文件 | 文件版本 | V2.5.0 |
| | 项目代号 | / | 页 码 | 第 0 页 共 1 页 |
| | 文件编号 | GL DM-56-24-en | | |

文件名称:

CALIBRATOR C5E0524 2024-06-10

适用范围: 1117 项目

编制: 刘嘉

审核: 刘金盈

批准: 郑剑通

日期: 2024.04.30

日期: 2024.04.30

日期: 2024.04.30

修 订 记 录

| 版本 | TCN/ECR/PCN 编号 | 修订内容概要 | 修改人 | 批准人 | 批准日期 |
|--------|-------------------|--------|-----|-----|------------|
| V2.5.0 | | 新建 | 刘嘉 | 郑剑通 | 2024.04.30 |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

发至:

总经办 质量部 财务部 营销部 客服 计划 采购 工程 仓库

设备研发 试剂研发 物料质量 设备生产 试剂生产 试剂质检 设备质检


由试剂研发


DM-CAL-5E

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR
 C5E0524

 2024-04-12

 2024-06-10

| Applicable Instruments | Parameter | Unit | Reference Value |
|---|-----------|--------------------|-----------------|
| DYMIND DH-800 series DH-800 CRP series DH-800 CS series (Technical File Version A1.3 or higher) | WBC | $\times 10^9/L$ | 9.20 |
| | RBC | $\times 10^{12}/L$ | 4.43 |
| | HGB | g/L | 134 |
| | MCV | fL | 93.7 |
| | PLT | $\times 10^9/L$ | 249 |

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.