

	文件类型	技术文件	文件版本	V2.5.0	
	项目代号	/	页 码	第 0 页 共 1 页	
	文件编号	GL DM-56-61			
文件名称： <p style="text-align: center;">CALIBRATOR CPF0524 2024-06-10</p> 适用范围：1117 项目					
编制：刘嘉 日期：2024.04.30		审核：刘金盈 日期：2024.04.30		批准：郑剑通 日期：2024.04.30	
修 订 记 录					
版本	TCN/ECR/PCN 编号	修订内容概要	修改人	批准人	批准日期
V2.5.0	/	新建	刘嘉	郑剑通	2024.04.30
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input type="checkbox"/> 营销部 <input type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input type="checkbox"/> 设备研发 <input type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input type="checkbox"/> 试剂质检 <input type="checkbox"/> 设备质检				
由试剂研发					

DM-CAL-PF

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT CPF0524

2024-04-12

2024-06-10

Applicable Instruments	Parameter	Unit	Reference Value
DYMIND DH-800 series DH-800 CRP series DH-800 CS series (Technical File Version A1.3 or higher)	PLT-F	$\times 10^9/L$	195

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.