

DM-RET

HEMATOLOGY CONTROL

CONTROL

📅 2024-04-12

🕒 2024-07-10

Applicable Instruments	Parameter	Unit	LOT RET0524L		LOT RET0524N		LOT RET0524H	
			Value	±	Value	±	Value	±
DYMIND	RET#	×10 ¹² /L	0.0433	±0.0217	0.2251	±0.1316	0.2312	±0.1156
DH-600	RET%	%	0.78	±0.39	4.88	±1.95	10.00	±4.00
DH-602								
DH-605	LFR	%	89.9	±27.0	89.8	±27.0	89.6	±27.0
DH-610								
DH-612	MFR	%	7.9	±12.0	8.3	±12.0	8.7	±13.0
DH-615								
(Technical File Version A2.0 or higher)	HFR	%	2.2	±3.0	1.9	±3.0	1.8	±3.0
	IRF	%	10.1	±15.0	10.2	±15.0	10.5	±16.0
	RHE	pg	16.3	±7.0	16.1	±6.0	15.9	±6.0

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0524L



RET0524N



RET0524H

DM-RET

HEMATOLOGY CONTROL

CONTROL

📅 2024-04-12

🕒 2024-07-10

Applicable Instruments	Parameter	Unit	LOT RET0524L		LOT RET0524N		LOT RET0524H	
DYMIND								
DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.4 or higher)	RET#	×10 ¹² /L	0.0361	±0.0288	0.2218	±0.1109	0.2034	±0.1017
	RET%	%	0.57	±0.46	4.26	±2.13	7.55	±3.78

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0524L



RET0524N



RET0524H