

DM-CAL

HEMATOLOGY CALIBRATOR

LOT N0624

2024-05-12

2024-07-10

Applicable Instruments	Parameter	Unit	Reference Value
DYMIND DP-H10、 DP-H12、 DP-H13、 DP-H15、 DP-H16、 DP-H17 (Technical File Version A3.0 or higher) ;	WBC	$\times 10^9/L$	9.46
	RBC	$\times 10^{12}/L$	4.51
	HGB	g/L	133
	MCV	fL	92.6
	PLT	$\times 10^9/L$	259

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.