
	文件类型	技术文档	文件版本	V3.2.0		
	项目代号	DM 3514	页 码	第 0 页 共 1 页		
	文件编号	GL DM-56-15-en				
文件名称： <p style="text-align: center;">Table of Serum Amyloid A Protein Calibrator Reference Value</p> 适用范围： <p style="text-align: center;">DM 3514 项目</p>						
编制：秦俊飞 日期：2024.06.13		审核：曾婷婷 日期：2024.06.13		批准：田文飞 日期：2024.06.13		
修 订 记 录						
版本	TCN/ECR/PCN 编号	修订内容概要		修改人	批准人	批准日期
V3.2.0	/	新增		秦俊飞	田文飞	2024.06.13
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input checked="" type="checkbox"/> 设备生产 <input checked="" type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由设转						

Table of Serum Amyloid A Protein Calibrator Reference Value

Serum Amyloid A Protein Calibrator

Code: DM3514

LOT 2024051801

 20250517

Model	Level	Reference Value (mg/L)
DH-100T、DH-102T、DH-103T、DH-105T、DH-106T、 DH-108T、DH-2000T[H1T1]、DH-2000T[H2T1]、 DH-2000T[H1T2]、DH-2000T[H2T2] (Applicable to technical document version A1.0 and above)	Calibrator 1	9.75
	Calibrator 2	20.40
	Calibrator 3	51.90
	Calibrator 4	105.75
	Calibrator 5	187.24
	Calibrator 6	240.30
	Calibrator 7	286.10
DH-800 CRP series DH-800 CS series (Technical File Version A1.3 or higher)	Calibrator 1	8.58
	Calibrator 2	20.80
	Calibrator 3	50.10
	Calibrator 4	104.50
	Calibrator 5	190.50
	Calibrator 6	218.00
	Calibrator 7	268.00