	文件类型	技术文件	文件版本	V10.8.0		
	项目代号	/	页码	第0页 共2页		
	文件编号	GL DM-56-05-en				
文件名称： <p style="text-align: center;">CALIBRATOR N0824 2024-09-10</p> 适用范围： 公司所有三分类仪器						
编制：刘嘉 日期：2024.07.18		审核：刘金盈 日期：2024.07.18		批准：郑剑通 日期：2024.07.18		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V10.8.0	/	新建		刘嘉	郑剑通	2024.07.18
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由试剂研发						


DM-CAL


HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT	N0824
-----	-------

	2024-07-13
---	------------

	2024-09-10
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31 Vet, DH36 Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	9.46
	RBC	$\times 10^{12}/L$	4.57
	HGB	g/L	132
	MCV	fL	91.1
	PLT	$\times 10^9/L$	235
DH21, DH23 DH25, DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.36
	RBC	$\times 10^{12}/L$	4.59
	HGB	g/L	133
	MCV	fL	85.0
	PLT	$\times 10^9/L$	243
DH20, DH22 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.45
	RBC	$\times 10^{12}/L$	4.59
	HGB	g/L	136
	MCV	fL	88.6
	PLT	$\times 10^9/L$	245
DH31 Vet, DH36 Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	9.45
	RBC	$\times 10^{12}/L$	4.58
	HGB	g/L	130
	MCV	fL	93.5
	PLT	$\times 10^9/L$	231

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.


DM-CAL


HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

 N0824

 2024-07-13

 2024-09-10

Applicable Instruments	Parameter	Unit	Reference Value
DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.71
	RBC	$\times 10^{12}/L$	4.55
	HGB	g/L	133
	MCV	fL	88.4
	PLT	$\times 10^9/L$	239
DP-H10 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.45
	RBC	$\times 10^{12}/L$	4.58
	HGB	g/L	140
	MCV	fL	93.5
	PLT	$\times 10^9/L$	231

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.