





# DM-RET

## HEMATOLOGY CONTROL

**Reference Values provided by DYMIND**
**CONTROL**
 2024-08-13

 2024-11-10

Applicable Instruments	Parameter	Unit	LOT RET0924L	LOT RET0924N	LOT RET0924H
DYMIND  DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	RET#	$\times 10^{12}/L$	0.0352 $\pm$ 0.0176	0.2237 $\pm$ 0.1316	0.2883 $\pm$ 0.1200
	RET%	%	0.64 $\pm$ 0.32	4.91 $\pm$ 1.96	12.27 $\pm$ 4.50
	LFR	%	90.0 $\pm$ 27.0	90.1 $\pm$ 27.0	89.7 $\pm$ 27.0
	MFR	%	7.7 $\pm$ 11.0	8.1 $\pm$ 12.0	8.2 $\pm$ 12.0
	HFR	%	2.3 $\pm$ 3.0	1.8 $\pm$ 3.0	2.1 $\pm$ 3.0
	IRF	%	10.0 $\pm$ 15.0	9.9 $\pm$ 15.0	10.3 $\pm$ 16.0
	RHE	pg	19.0 $\pm$ 8.0	17.0 $\pm$ 7.0	16.5 $\pm$ 7.0

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0924L




RET0924N




RET0924H

# DM-RET

## HEMATOLOGY CONTROL

**Reference Values provided by DYMIND**
**CONTROL**
 2024-08-13

 2024-11-10

Applicable Instruments	Parameter	Unit	LOT RET0924L	LOT RET0924N	LOT RET0924H
DYMIND DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.4 or higher)	RET#	$\times 10^{12}/L$	0.0113 ±0.0090	0.1753 ±0.0877	0.1527 ±0.0763
	RET%	%	0.21 ±0.17	3.94 ±1.97	6.51 ±3.26

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0924L



RET0924N



RET0924H