
	文件类型	技术文档	文件版本	V3.5.0		
	项目代号	DM 3514	页 码	第 0 页 共 1 页		
	文件编号	GL DM-56-31-en				
文件名称： <p style="text-align: center;">Table of Serum Amyloid A Protein Calibrator Reference Value</p> 适用范围： <p style="text-align: center;">DM 3514 项目</p>						
编制：刘嘉 日期：2024.09.05		审核：刘金盈 日期：2024.09.05		批准：郑剑通 日期：2024.09.05		
修 订 记 录						
版本	TCN/ECR/PCN 编号	修订内容概要		修改人	批准人	批准日期
V3.5.0	/	新增		刘嘉	郑剑通	2024.09.05
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input checked="" type="checkbox"/> 设备生产 <input checked="" type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由试剂研发						

Table of Serum Amyloid A Protein Calibrator Reference Value

Serum Amyloid A Protein Calibrator

Code: DM3514

LOT 2024082701

 20250826

Applicable Instruments	Level	Reference Value (mg/L)
DH-2000T series DH-100T、DH-102T DH-103T、DH-105T DH-106T、DH-108T (Applicable to technical document version A2.0 or higher)	Calibrator 1	10.50
	Calibrator 2	18.80
	Calibrator 3	48.50
	Calibrator 4	98.50
	Calibrator 5	209.50
	Calibrator 6	271.20
	Calibrator 7	289.00
DH-800 CS series (Applicable to technical document version A1.3 or higher)	Calibrator 1	8.80
	Calibrator 2	17.40
	Calibrator 3	46.80
	Calibrator 4	99.80
	Calibrator 5	196.87
	Calibrator 6	265.97
	Calibrator 7	292.00