
	文件类型	技术文件	文件版本	V5.6.0	
	项目代号	DM3113	页 码	第 0 页 共 3 页	
	文件编号	GL DM-56-09-en			
文件名称: <p style="text-align: center;">Table of CRP Calibrator Reference Value</p>					
适用范围: <p style="text-align: center;">DM3113 项目</p>					
编制: 秦俊飞 日期: 2024.09.26		审核: 曾婷婷 日期: 2024.09.26		批准: 田文飞 日期: 2024.09.26	
修 订 记 录					
版本	TCN/ECR/P CN 编号	修订内容概要	修改人	批准人	批准日期
V5.6.0	/	新建	秦俊飞	田文飞	2024.09.26
发至:	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input type="checkbox"/> 营销部 <input type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input type="checkbox"/> 设备研发 <input type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input type="checkbox"/> 试剂质检 <input type="checkbox"/> 设备质检				
由试剂研发					

Table of CRP Calibrator Reference Value

C-Reactive Protein Calibrator

Code: DM3113

 2024083001

 20250829

Model	Level (multi-point)	Reference Value (mg/L)
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.10 and above with heating)	Calibrator 1	5.9
	Calibrator 2	19.5
	Calibrator 3	49.4
	Calibrator 4	107.1
	Calibrator 5	149.4
	Calibrator 6	179.4
DH51CRP, DH53CRP, DH56CRP D1-CRP, D3-CRP, D5-CRP (Applicable to technical document version A11.10 and above without heating)	Calibrator 1	6.2
	Calibrator 2	19.9
	Calibrator 3	50.9
	Calibrator 4	107.0
	Calibrator 5	152.3
	Calibrator 6	197.1
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A5.11 and below)	Calibrator 1	4.6
	Calibrator 2	17.3
	Calibrator 3	45.8
	Calibrator 4	92.8
	Calibrator 5	114.5
	Calibrator 6	146.5
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A6.0 and above with heating)	Calibrator 1	5.9
	Calibrator 2	23.6
	Calibrator 3	53.8
	Calibrator 4	94.3
	Calibrator 5	130.8
	Calibrator 6	176.3
DH71CRP, DH73CRP D2-CRP, D7-CRP (Applicable to technical document versions A6.0 and above without heating)	Calibrator 1	6.2
	Calibrator 2	20.4
	Calibrator 3	51.8
	Calibrator 4	109.1
	Calibrator 5	134.3
	Calibrator 6	167.5

D6-CRP, DH76CRP (Applicable to technical document version B2.0)	Calibrator 1	5.1
	Calibrator 2	19.1
	Calibrator 3	54.3
	Calibrator 4	104.2
	Calibrator 5	138.8
	Calibrator 6	173.5
D6-CRP, DH76CRP (Applicable to technical document version B2.2 and above)	Calibrator 1	5.5
	Calibrator 2	19.9
	Calibrator 3	52.2
	Calibrator 4	107.2
	Calibrator 5	134.5
	Calibrator 6	173.2
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Applicable to technical document versions A8.1 and below with heating)	Calibrator 1	6.3
	Calibrator 2	19.9
	Calibrator 3	48.5
	Calibrator 4	107.2
	Calibrator 5	145.5
	Calibrator 6	187.0
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Applicable to technical document versions A8.1 and below without heating)	Calibrator 1	6.4
	Calibrator 2	20.3
	Calibrator 3	54.2
	Calibrator 4	95.6
	Calibrator 5	142.3
	Calibrator 6	184.0
DM79X, DM78X, DM72X (Applicable to technical document version A1.0 and above)	Calibrator 1	5.5
	Calibrator 2	21.7
	Calibrator 3	52.7
	Calibrator 4	102.8
	Calibrator 5	130.8
	Calibrator 6	176.8
DH36X, DH33X, DH31X (Applicable to technical document versions A1.0 and above)	Calibrator 1	5.2
	Calibrator 2	18.4
	Calibrator 3	49.7
	Calibrator 4	98.8
	Calibrator 5	139.5
	Calibrator 6	174.7

DP-H10、DP-H12、DP-H13、DP-H15、DP-H16、 DP-H17 (Applicable to technical document versions A1.0 and above)	Calibrator 1	7.2
	Calibrator 2	21.3
	Calibrator 3	54.5
	Calibrator 4	104.0
	Calibrator 5	149.4
	Calibrator 6	174.9
DH-800 CRP series DH-800 CS series (Technical File Version A1.3 or higher)	Calibrator 1	4.75
	Calibrator 2	18.90
	Calibrator 3	45.01
	Calibrator 4	80.90
	Calibrator 5	94.21
	Calibrator 6	136.50