



DM-CAL

HEMATOLOGY CALIBRATOR

LOT N1024

 2024-09-12 2024-11-10

Applicable Instruments	Parameter	Unit	Reference Value
DYMIND DP-H10、 DP-H12、 DP-H13、 DP-H15、 DP-H16、 DP-H17 (Technical File Version A3.0 or higher) ;	WBC	$\times 10^9/L$	9.83
	RBC	$\times 10^{12}/L$	4.67
	HGB	g/L	130
	MCV	fL	88.4
	PLT	$\times 10^9/L$	236

【NOTE】

1. The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.