
	文件类型	技术文件	模板版本	V4.5.0		
	项目代号	/	页 码	第 0 页 共 1 页		
	文件编号	GL DM-56-46-en				
文件名称： <h2 style="text-align: center;">Coagulation Control Reference Value Table 2026-07-04</h2> 适用范围：9102 项目						
编制：陈贞竹 日期：2024.10.18		审核：曹佳强 日期：2024.10.18		批准：曹佳强 日期：2024.10.18		
修 订 记 录						
版本	TCN/ECR/PCN 编号	修订内容概要		修改人	批准人	批准日期
V4.5.0	/	新建		陈贞竹	曹佳强	2024.10.18
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由试剂研发						


Coagulation Control Reference Value Table

Coagulation Control-level 2

CONTROL

LOT	2024070501
-----	------------

	20240705
---	----------

	20260704
---	----------

Applicable Instrument	Parameter	Unit	Principle	Reference Value	Deviation Limit
CA1200	PT	s	Optical method	36.6	± 7.3
		s	Magnetic bead	38.5	± 7.7
CA1203	APTT	s	Optical method	44.0	± 8.8
CA1205		s	Magnetic bead	43.0	± 8.6
DC1210	FIB	g/L	Optical method	1.28	± 0.26
DC1250		g/L	Magnetic bead	1.31	± 0.26
DC1280	AT-III	%	Optical method	55.5	± 11.1

【NOTE】

1. The calibrator should be stored in the refrigerator at 2°C - 8°C. After reconstitution, the calibrator can be stable for 4 hours at room temperature (10°C - 30°C), and can be stable for 24 hours at 2°C - 8°C.
2. The calibrator cannot be frozen after reconstitution.
3. The calibrator is for IVD only.