



	文件类型	技术文件	文件版本	V4.11.0		
	项目代号	/	页码	第0页 共2页		
	文件编号	GL DM-56-17-en				
文件名称： <p style="text-align: center;"><b>CONTROL DM-RET RET1124 2025-01-10</b></p> 适用范围： 1116项目、7003项目						
编制：薛泽昊 日期：2024.11.06		审核：刘金盈 日期：2024.11.06		批准：郑剑通 日期：2024.11.06		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V4.11.0	/	新建		薛泽昊	郑剑通	2024.11.06
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由试剂研发						

# DM-RET

## HEMATOLOGY CONTROL

**Reference Values provided by DYMIND**
**CONTROL**
 2024-10-13

 2025-01-10

Applicable Instruments	Parameter	Unit	LOT RET1124L	LOT RET1124N	LOT RET1124H
DYMIND  DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	RET#	$\times 10^{12}/L$	0.0401 $\pm 0.0200$	0.2283 $\pm 0.1316$	0.2390 $\pm 0.1195$
	RET%	%	0.71 $\pm 0.36$	5.08 $\pm 2.03$	10.13 $\pm 4.05$
	LFR	%	89.9 $\pm 27.0$	89.7 $\pm 27.0$	89.3 $\pm 27.0$
	MFR	%	8.2 $\pm 12.0$	8.4 $\pm 13.0$	8.6 $\pm 13.0$
	HFR	%	1.9 $\pm 3.0$	1.9 $\pm 3.0$	2.1 $\pm 3.0$
	IRF	%	10.2 $\pm 15.0$	10.3 $\pm 15.0$	10.7 $\pm 16.0$
	RHE	pg	16.1 $\pm 6.0$	15.6 $\pm 6.0$	16.8 $\pm 7.0$

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET1124L



RET1124N




RET1124H


# DM-RET

## HEMATOLOGY CONTROL

Reference Values provided by DYMIND

**CONTROL**

 2024-10-13

 2025-01-10

Applicable Instruments	Parameter	Unit	LOT RET1124L	LOT RET1124N	LOT RET1124H
DYMIND DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.4 or higher)	RET#	$\times 10^{12}/L$	0.0201 $\pm 0.0161$	0.1683 $\pm 0.0841$	0.1099 $\pm 0.0549$
	RET%	%	0.37 $\pm 0.29$	3.84 $\pm 1.92$	4.68 $\pm 2.34$

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET1124L



RET1124N



RET1124H