
	文件类型	技术文件	文件版本	V5.7.0		
	项目代号	/	页码	第0页 共1页		
	文件编号	GL DM-56-06-en				
文件名称： <p style="text-align: center;"><b>Table of C-Reactive Protein Control (CRP) Reference Value</b></p> 适用范围： DM3113项目、DM3107项目、DM3513项目						
编制：刘嘉 日期：2024.11.25		审核：刘金盈 日期：2024.11.25		批准：郑剑通 日期：2024.11.25		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V5.7.0	/	新建		刘嘉	郑剑通	2024.11.25
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由试剂研发						

# Table of C-Reactive Protein Control (CRP) Reference Value

## C-Reactive Protein Control

LOT 2024100801


 2025-10-07

Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)	Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)
DH51CRP, DH53CRP, DH56CRP, D1-CRP, D3-CRP, D5-CRP (Technical file version A11.10 or higher with heating)	Control 1	<b>4.1</b>	±1.2	DH51CRP, DH53CRP, DH56CRP, D1-CRP, D3-CRP, D5-CRP (Technical file version A11.10 or higher without heating)	Control 1	<b>4.0</b>	±1.2
	Control 2	<b>17.0</b>	±5.1		Control 2	<b>17.3</b>	±5.2
	Control 3	<b>42.3</b>	±12.7		Control 3	<b>46.6</b>	±14.0
DH71CRP, DH73CRP, D2-CRP, D7-CRP (Technical file version A6.0 or higher with heating)	Control 1	<b>3.3</b>	±1.0	DH71CRP, DH73CRP, D2-CRP, D7-CRP (Technical file version A6.0 or higher without heating)	Control 1	<b>3.6</b>	±1.1
	Control 2	<b>16.4</b>	±4.9		Control 2	<b>16.8</b>	±5.0
	Control 3	<b>43.9</b>	±13.2		Control 3	<b>43.9</b>	±13.2
D6-CRP, DH76CRP (Technical file version B2.0)	Control 1	<b>3.6</b>	±1.1	D6-CRP, DH76CRP (Technical file version B2.2 or higher)	Control 1	<b>3.4</b>	±1.0
	Control 2	<b>17.7</b>	±5.3		Control 2	<b>16.7</b>	±5.0
	Control 3	<b>44.3</b>	±13.3		Control 3	<b>43.8</b>	±13.1
DF1-CRP, DF3-CRP, DF5-CRP, DF50CRP, DF52CRP, DF53CRP (Technical file version A6.0 or higher with heating)	Control 1	<b>3.4</b>	±1.0	DF1-CRP, DF3-CRP, DF5-CRP, DF50CRP, DF52CRP, DF53CRP (Technical file versions A8.1 and below without heating)	Control 1	<b>3.7</b>	±1.1
	Control 2	<b>16.3</b>	±4.9		Control 2	<b>18.3</b>	±5.5
	Control 3	<b>48.6</b>	±14.6		Control 3	<b>49.4</b>	±14.8

# Table of C-Reactive Protein Control (CRP) Reference Value

## C-Reactive Protein Control

LOT 2024100801

 2025-10-07

Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)	Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)
DM79X, DM78X, DM72X (Technical file version A1.0 or higher )	Control 1	<b>3.5</b>	±1.1	DH36X, DH33X, DH31X (Technical file versions A1.0 or higher )	Control 1	<b>3.3</b>	±1.0
	Control 2	<b>16.9</b>	±5.1		Control 2	<b>16.5</b>	±5.0
	Control 3	<b>43.4</b>	±13.0		Control 3	<b>42.1</b>	±12.6
DP-H10, DP-H12, DP-H13, DP-H15, DP-H16, DP-H17 (Technical file versions A1.0 or higher)	Control 1	<b>3.7</b>	±1.1	DH-800 CRP series DH-800 CS series (Technical file Version A1.3 or higher )	Control 1	<b>3.19</b>	±0.96
	Control 2	<b>16.6</b>	±5.0		Control 2	<b>16.07</b>	±4.82
	Control 3	<b>44.8</b>	±13.4		Control 3	<b>39.50</b>	±11.85
DH-2000T series , DH-100T, DH-102T DH-103T, DH-105T DH-106T, DH-108T (Technical file version A2.0 or higher)	Control 1	<b>3.14</b>	±0.94	/	Control 1	/	/
	Control 2	<b>16.50</b>	±4.95		Control 2	/	/
	Control 3	<b>41.92</b>	±12.58		Control 3	/	/