



	文件类型	技术文件	文件版本	V5.1.0		
	项目代号	/	页码	第0页 共2页		
	文件编号	GL DM-56-17-en				
文件名称： <p style="text-align: center;">CONTROL DM-RET RET0125 2025-03-10</p> 适用范围： 1116项目、7003项目						
编制：郭罕星 日期：2024.12.19		审核：刘金盈 日期：2024.12.19		批准：郑剑通 日期：2024.12.19		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V5.1.0	/	新建		郭罕星	郑剑通	2024.12.19
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由试剂研发						

DM-RET

HEMATOLOGY CONTROL

Reference Values provided by DYMIND
CONTROL
 2024-12-11

 2025-03-10

Applicable Instruments	Parameter	Unit	LOT RET0125L	LOT RET0125N	LOT RET0125H
DYMIND DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	RET#	$\times 10^{12}/L$	0.0376 ± 0.0188	0.2202 ± 0.1316	0.2436 ± 0.1200
	RET%	%	0.69 ± 0.35	4.86 ± 1.94	10.25 ± 4.10
	LFR	%	89.8 ± 27.0	89.9 ± 27.0	89.6 ± 27.0
	MFR	%	8.1 ± 12.0	8.1 ± 12.0	8.5 ± 13.0
	HFR	%	2.1 ± 3.0	2.0 ± 3.0	1.9 ± 3.0
	IRF	%	10.2 ± 15.0	10.1 ± 15.0	10.4 ± 16.0
	RHE	pg	16.3 ± 7.0	13.7 ± 5.0	15.0 ± 6.0

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0125L




RET0125N




RET0125H

DM-RET

HEMATOLOGY CONTROL

Reference Values provided by DYMIND
CONTROL
 2024-12-11

 2025-03-10

Applicable Instruments	Parameter	Unit	LOT RET0125L	LOT RET0125N	LOT RET0125H
DYMIND DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.4 or higher)	RET#	$\times 10^{12}/L$	0.0124 ± 0.0099	0.1691 ± 0.0846	0.1744 ± 0.0872
	RET%	%	0.23 ± 0.18	3.81 ± 1.91	7.35 ± 3.68

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0125L



RET0125N



RET0125H