	文件类型	技术文件	文件版本	V11.2.0	
	项目代号	/	页码	第0页 共5页	
	文件编号	GL DM-56-02-en			
文件名称： <p style="text-align: center;">CALIBRATOR NK0225 2025-03-10</p> 适用范围： 公司所有五分类仪器					
编制：刘嘉 日期：2025.01.20		审核：刘金盈 日期：2025.01.20		批准：程军 日期：2025.01.20	
修 订 记 录					
版本	TCN/ECR/PCN编号	修订内容概要	修改人	批准人	批准日期
V11.2.0	/	新建	刘嘉	程军	2025.01.20
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由试剂研发					

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
LOT NK0225

M 2025-01-10

H 2025-03-10



Applicable Instruments	Parameter	Unit	Reference Value
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A9.0 to A11.9)	WBC	$\times 10^9/L$	9.48
	RBC	$\times 10^{12}/L$	4.40
	HGB	g/L	130
	MCV	fL	92.9
	PLT	$\times 10^9/L$	250
D1-CRP, D3-CRP, D5-CRP DH51CRP, DH53CRP, DH56CRP (Technical File Version A11.11 or higher)	WBC	$\times 10^9/L$	10.35
	RBC	$\times 10^{12}/L$	4.45
	HGB	g/L	129
	MCV	fL	87.9
	PLT	$\times 10^9/L$	239
UN71, UN73, UN76, DH73 (Technical File Version A5.5 to A6.2)	WBC	$\times 10^9/L$	9.39
	RBC	$\times 10^{12}/L$	4.42
	HGB	g/L	133
	MCV	fL	89.6
	PLT	$\times 10^9/L$	244
UN71, UN73, UN76, DH73 (Technical File Version A6.3 or higher)	WBC	$\times 10^9/L$	9.88
	RBC	$\times 10^{12}/L$	4.50
	HGB	g/L	134
	MCV	fL	89.5
	PLT	$\times 10^9/L$	239

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **NK0225**
 **2025-01-10**
 **2025-03-10**



Applicable Instruments	Parameter	Unit	Reference Value
DH73 Vet (Technical File Version B5.5 or higher)	RBC	$\times 10^{12}/L$	4.61
	HGB	g/L	135
	MCV	fL	92.6
	PLT	$\times 10^9/L$	244
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A11.2 to A11.6)	WBC	$\times 10^9/L$	9.13
	RBC	$\times 10^{12}/L$	4.48
	HGB	g/L	131
	MCV	fL	91.3
	PLT	$\times 10^9/L$	248
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version A12.0 or higher)	WBC	$\times 10^9/L$	9.48
	RBC	$\times 10^{12}/L$	4.54
	HGB	g/L	132
	MCV	fL	87.9
	PLT	$\times 10^9/L$	243
DF50, DF51, DF52 DF53, DF55, DF56 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.45
	RBC	$\times 10^{12}/L$	4.43
	HGB	g/L	133
	MCV	fL	88.9
	PLT	$\times 10^9/L$	237

【NOTE】

1. The calibrator should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 7 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **NK0225**
 **2025-01-10**
 **2025-03-10**



Applicable Instruments	Parameter	Unit	Reference Value
DH71, DH76 (Technical File Version A4.1 to A10.3) DH51, DH53, DH56 (Technical File Version A8.1 or higher)	WBC	$\times 10^9/L$	9.75
	RBC	$\times 10^{12}/L$	4.26
	HGB	g/L	132
	MCV	fL	92.9
	PLT	$\times 10^9/L$	229
DH71, DH76 (Technical File Version A10.4)	WBC	$\times 10^9/L$	8.87
	RBC	$\times 10^{12}/L$	4.53
	HGB	g/L	130
	MCV	fL	88.7
	PLT	$\times 10^9/L$	240
DH71, DH76 (Technical File Version A10.5 or higher)	WBC	$\times 10^9/L$	10.19
	RBC	$\times 10^{12}/L$	4.59
	HGB	g/L	135
	MCV	fL	88.3
	PLT	$\times 10^9/L$	240
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.0 to A6.5)	WBC	$\times 10^9/L$	9.75
	RBC	$\times 10^{12}/L$	4.48
	HGB	g/L	134
	MCV	fL	88.5
	PLT	$\times 10^9/L$	241

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **NK0225**
 **2025-01-10**
 **2025-03-10**



Applicable Instruments	Parameter	Unit	Reference Value
D2-CRP, D7-CRP DH71CRP, DH73CRP (Technical File Version A6.6 or higher)	WBC	$\times 10^9/L$	10.46
	RBC	$\times 10^{12}/L$	4.50
	HGB	g/L	135
	MCV	fL	87.8
	PLT	$\times 10^9/L$	237
D6-CRP, DH76CRP (Technical File Version B1.0 to B2.0)	WBC	$\times 10^9/L$	9.60
	RBC	$\times 10^{12}/L$	4.44
	HGB	g/L	134
	MCV	fL	88.2
	PLT	$\times 10^9/L$	243
D6-CRP, DH76CRP (Technical File Version B2.2 or higher)	WBC	$\times 10^9/L$	10.23
	RBC	$\times 10^{12}/L$	4.59
	HGB	g/L	135
	MCV	fL	88.6
	PLT	$\times 10^9/L$	244
DF50 Vet, DF52 Vet DF55 Vet, DF56 Vet (Technical File Version A8.0 or higher and B1.0 or higher)	WBC	$\times 10^9/L$	9.43
	RBC	$\times 10^{12}/L$	4.60
	HGB	g/L	136
	MCV	fL	90.5
	PLT	$\times 10^9/L$	245

【NOTE】

- The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
- Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
- Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.

DM-CAL PLUS

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND
CALIBRATOR
 **NK0225**
 **2025-01-10**
 **2025-03-10**

Applicable Instruments	Parameter	Unit	Reference Value
DF1-CRP, DF3-CRP, DF5-CRP DF50CRP, DF52CRP, DF53CRP (Technical File Version A6.1 or higher)	WBC	$\times 10^9/L$	10.30
	RBC	$\times 10^{12}/L$	4.36
	HGB	g/L	131
	MCV	fL	87.8
	PLT	$\times 10^9/L$	242
DM71X, DM72X DM78X, DM79X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.75
	RBC	$\times 10^{12}/L$	4.47
	HGB	g/L	134
	MCV	fL	88.2
	PLT	$\times 10^9/L$	240
UN71 Vet, UN73 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.55
	RBC	$\times 10^{12}/L$	4.51
	HGB	g/L	133
	MCV	fL	88.6
	PLT	$\times 10^9/L$	233
DM60 Vet, DM61 Vet DM62 Vet, DM63 Vet (Technical File Version A1.4 or higher)	WBC	$\times 10^9/L$	8.54
	RBC	$\times 10^{12}/L$	4.23
	HGB	g/L	128
	MCV	fL	87.7
	PLT	$\times 10^9/L$	234

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.