
	文件类型	技术文件	文件版本	V5.8.0	
	项目代号	/	页码	第0页 共2页	
	文件编号	GL DM-56-06-en			
文件名称： <p style="text-align: center;">Table of C-Reactive Protein Control (CRP) Reference Value</p> 适用范围： DM3113项目、DM3513项目					
编制：石丽岐 日期：2025.01.04		审核：刘金盈 日期：2025.01.04		批准：郑剑通 日期：2025.01.04	
修 订 记 录					
版本	TCN/ECR/PCN编号	修订内容概要	修改人	批准人	批准日期
V5.8.0	/	新建	石丽岐	郑剑通	2025.01.04
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由试剂研发					

Table of C-Reactive Protein Control (CRP) Reference Value

C-Reactive Protein Control

LOT 2024111501


 2025-11-14

Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)	Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)
DH51CRP, DH53CRP, DH56CRP, D1-CRP, D3-CRP, D5-CRP (Technical file version A11.10 or higher with heating)	Control 1	3.3	±1.0	DH51CRP, DH53CRP, DH56CRP, D1-CRP, D3-CRP, D5-CRP (Technical file version A11.10 or higher without heating)	Control 1	3.7	±1.1
	Control 2	16.2	±4.9		Control 2	17.4	±5.2
	Control 3	44.5	±13.3		Control 3	47.9	±14.4
DH71CRP, DH73CRP, D2-CRP, D7-CRP (Technical file version A6.0 or higher with heating)	Control 1	3.4	±1.0	DH71CRP, DH73CRP, D2-CRP, D7-CRP (Technical file version A6.0 or higher without heating)	Control 1	3.3	±1.0
	Control 2	17.9	±5.4		Control 2	16.9	±5.1
	Control 3	49.2	±14.8		Control 3	44.5	±13.3
D6-CRP, DH76CRP (Technical file version B2.0)	Control 1	3.4	±1.0	D6-CRP, DH76CRP (Technical file version B2.2 or higher)	Control 1	3.2	±1.0
	Control 2	16.6	±5.0		Control 2	16.2	±4.8
	Control 3	44.4	±13.3		Control 3	47.0	±14.1
DF1-CRP, DF3-CRP, DF5-CRP, DF50CRP, DF52CRP, DF53CRP (Technical file version A6.0 or higher with heating)	Control 1	3.3	±1.0	DF1-CRP, DF3-CRP, DF5-CRP, DF50CRP, DF52CRP, DF53CRP (Technical file versions A8.1 and below without heating)	Control 1	3.8	±1.1
	Control 2	15.9	±4.8		Control 2	17.6	±5.3
	Control 3	45.7	±13.7		Control 3	44.6	±13.4

Table of C-Reactive Protein Control (CRP) Reference Value

C-Reactive Protein Control

LOT 2024111501

 2025-11-14

Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)	Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)
DM79X, DM78X, DM72X (Technical file version A1.0 or higher)	Control 1	3.4	±1.0	DH36X, DH33X, DH31X (Technical file versions A1.0 or higher)	Control 1	3.5	±1.0
	Control 2	17.1	±5.1		Control 2	15.8	±4.7
	Control 3	41.7	±12.5		Control 3	45.1	±13.5
DP-H10, DP-H12, DP-H13, DP-H15, DP-H16, DP-H17 (Technical file versions A1.0 or higher)	Control 1	4.0	±1.2	DH-800 CRP series DH-800 CS series (Technical file Version A1.3 or higher)	Control 1	3.45	±1.04
	Control 2	19.1	±5.7		Control 2	15.80	±4.74
	Control 3	45.7	±13.7		Control 3	38.99	±11.70
DH-2000T series , DH-100T, DH-102T DH-103T, DH-105T DH-106T, DH-108T (Technical file version A2.0 or higher)	Control 1	3.24	±0.97	/	Control 1	/	/
	Control 2	16.92	±5.08		Control 2	/	/
	Control 3	43.30	±12.99		Control 3	/	/