


Table of Serum Amyloid A Protein Calibrator (SAA) Reference Value

Serum Amyloid A Protein Calibrator

LOT 2025012201

 2026-01-21

Applicable Instruments	DM79X、DM78X、DM72X (Applicable to technical file version A1.0 or higher)	DP-H10、DP-H12、DP-H13、DP-H15、DP-H16、 DP-H17 (Applicable to technical file version A1.0 or higher)
Level (multi-point)	Reference Value (mg/L)	Reference Value (mg/L)
Calibrator 1	11.08	7.9
Calibrator 2	22.06	18.9
Calibrator 3	54.05	52.1
Calibrator 4	102.24	74.5
Calibrator 5	175.77	149.5
Calibrator 6	201.60	/
Calibrator 7	278.33	/