




# DM-CAL


## HEMATOLOGY CALIBRATOR

**Reference Values provided by DYMIND**

<b>CALIBRATOR</b>
-------------------

LOT	N0325
-----	-------

	2025-02-10
---	------------

	2025-04-10
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31 Vet, DH36 Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	9.09
	RBC	$\times 10^{12}/L$	4.47
	HGB	g/L	131
	MCV	fL	94.2
	PLT	$\times 10^9/L$	236
DH21, DH23 DH25, DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	8.84
	RBC	$\times 10^{12}/L$	4.41
	HGB	g/L	128
	MCV	fL	85.6
	PLT	$\times 10^9/L$	237
DH20, DH22 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.31
	RBC	$\times 10^{12}/L$	4.51
	HGB	g/L	132
	MCV	fL	91.1
	PLT	$\times 10^9/L$	244
DH31 Vet, DH36 Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	8.88
	RBC	$\times 10^{12}/L$	4.39
	HGB	g/L	124
	MCV	fL	93.8
	PLT	$\times 10^9/L$	230

**【NOTE】**

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.


# DM-CAL


## HEMATOLOGY CALIBRATOR

**Reference Values provided by DYMIND**

<b>CALIBRATOR</b>
-------------------

LOT	N0325
-----	-------

	2025-02-10
---	------------

	2025-04-10
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.16
	RBC	$\times 10^{12}/L$	4.40
	HGB	g/L	128
	MCV	fL	89.0
	PLT	$\times 10^9/L$	234
DP-H10 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	8.88
	RBC	$\times 10^{12}/L$	4.39
	HGB	g/L	134
	MCV	fL	93.8
	PLT	$\times 10^9/L$	230

**【NOTE】**

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.