



	文件类型	技术文件	文件版本	V5.2.0		
	项目代号	/	页码	第0页 共2页		
	文件编号	GL DM-56-17-en				
文件名称： <p style="text-align: center;"><b>CONTROL DM-RET RET0325 2025-05-10</b></p> 适用范围： 1116项目、7003项目						
编制：刘嘉 日期：2025.03.05		审核：刘金盈 日期：2025.03.05		批准：程军 日期：2025.03.05		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V5.2.0	/	新建		刘嘉	程军	2025.03.05
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由试剂研发						

# DM-RET

## HEMATOLOGY CONTROL

**Reference Values provided by DYMIND**
**CONTROL**
 2025-02-10

 2025-05-10

Applicable Instruments	Parameter	Unit	<span style="border: 1px solid black; padding: 1px;">LOT</span> RET0325L	<span style="border: 1px solid black; padding: 1px;">LOT</span> RET0325N	<span style="border: 1px solid black; padding: 1px;">LOT</span> RET0325H
DYMIND	RET#	$\times 10^{12}/L$	0.0329 $\pm 0.0165$	0.2121 $\pm 0.1273$	0.3179 $\pm 0.1200$
DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	RET%	%	0.59 $\pm 0.30$	4.73 $\pm 1.89$	13.25 $\pm 4.50$
	LFR	%	89.0 $\pm 27.0$	89.8 $\pm 27.0$	89.6 $\pm 27.0$
	MFR	%	8.5 $\pm 13.0$	8.3 $\pm 12.0$	8.4 $\pm 13.0$
	HFR	%	2.5 $\pm 4.0$	1.9 $\pm 3.0$	2.0 $\pm 3.0$
	IRF	%	11.0 $\pm 17.0$	10.2 $\pm 15.0$	10.4 $\pm 16.0$
	RHE	pg	17.8 $\pm 7.0$	14.7 $\pm 6.0$	14.5 $\pm 6.0$

**【NOTE】**

1. The controls should be stored in refrigerator (2℃~8℃). After opening, it will keep stable for 14 days when it is stored airtight at 2℃~8℃.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15℃~30℃).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0325L




RET0325N




RET0325H

# DM-RET

## HEMATOLOGY CONTROL

**Reference Values provided by DYMIND**
**CONTROL**
 2025-02-10

 2025-05-10

Applicable Instruments	Parameter	Unit	LOT RET0325L	LOT RET0325N	LOT RET0325H
DYMIND  DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.4 or higher)	RET#	$\times 10^{12}/L$	0.0140 $\pm 0.0112$	0.1634 $\pm 0.0817$	0.1762 $\pm 0.0881$
	RET%	%	0.26 $\pm 0.21$	3.78 $\pm 1.89$	7.40 $\pm 3.70$

**【NOTE】**

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0325L



RET0325N



RET0325H