
	文件类型	技术文件	文件版本	V5.9.0		
	项目代号	/	页码	第0页 共2页		
	文件编号	GL DM-56-06-en				
文件名称： <p style="text-align: center;"><b>Table of C-Reactive Protein Control (CRP) Reference Value 2024123001</b></p> 适用范围： DM3113项目、DM3513项目						
编制：刘嘉 日期：2025.03.07		审核：刘金盈 日期：2025.03.07		批准：程军 日期：2025.03.07		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V5.9.0	/	新建		刘嘉	程军	2025.03.07
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库  <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由试剂研发						

# Table of C-Reactive Protein Control (CRP) Reference Value

## C-Reactive Protein Control

LOT 2024123001


 2025-12-29

Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)	Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)
DH51CRP, DH53CRP, DH56CRP, D1-CRP, D3-CRP, D5-CRP (Technical file version A11.10 or higher with heating)	Control 1	<b>4.3</b>	±1.3	DH51CRP, DH53CRP, DH56CRP, D1-CRP, D3-CRP, D5-CRP (Technical file version A11.10 or higher without heating)	Control 1	<b>4.2</b>	±1.3
	Control 2	<b>21.9</b>	±6.6		Control 2	<b>21.8</b>	±6.5
	Control 3	<b>48.1</b>	±14.4		Control 3	<b>50.0</b>	±15.0
DH71CRP, DH73CRP, D2-CRP, D7-CRP (Technical file version A6.0 or higher with heating)	Control 1	<b>3.0</b>	±0.9	DH71CRP, DH73CRP, D2-CRP, D7-CRP (Technical file version A6.0 or higher without heating)	Control 1	<b>3.9</b>	±1.2
	Control 2	<b>18.4</b>	±5.5		Control 2	<b>19.6</b>	±5.9
	Control 3	<b>46.1</b>	±13.8		Control 3	<b>47.7</b>	±14.3
D6-CRP, DH76CRP (Technical file version B2.0)	Control 1	<b>3.8</b>	±1.1	D6-CRP, DH76CRP (Technical file version B2.2 or higher)	Control 1	<b>3.8</b>	±1.1
	Control 2	<b>20.0</b>	±6.0		Control 2	<b>19.6</b>	±5.9
	Control 3	<b>48.2</b>	±14.5		Control 3	<b>46.1</b>	±13.8
DF1-CRP, DF3-CRP, DF5-CRP, DF50CRP, DF52CRP, DF53CRP (Technical file version A6.0 or higher with heating)	Control 1	<b>4.3</b>	±1.3	DF1-CRP, DF3-CRP, DF5-CRP, DF50CRP, DF52CRP, DF53CRP (Technical file versions A8.1 and below without heating)	Control 1	<b>4.2</b>	±1.3
	Control 2	<b>21.2</b>	±6.4		Control 2	<b>20.9</b>	±6.3
	Control 3	<b>51.5</b>	±15.5		Control 3	<b>51.2</b>	±15.4

# Table of C-Reactive Protein Control (CRP) Reference Value

## C-Reactive Protein Control

LOT 2024123001

 2025-12-29

Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)	Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)
DM79X, DM78X, DM72X (Technical file version A1.0 or higher)	Control 1	<b>3.77</b>	±1.1	DH36X, DH33X, DH31X (Technical file versions A1.0 or higher)	Control 1	<b>3.7</b>	±1.1
	Control 2	<b>19.69</b>	±5.9		Control 2	<b>20.0</b>	±6.0
	Control 3	<b>44.33</b>	±13.3		Control 3	<b>47.6</b>	±14.3
DP-H10, DP-H12, DP-H13, DP-H15, DP-H16, DP-H17 (Technical file versions A1.0 or higher)	Control 1	<b>4.1</b>	±1.2	DH-800 CRP series DH-800 CS series (Technical file Version A1.3 or higher)	Control 1	<b>3.55</b>	±1.07
	Control 2	<b>18.8</b>	±5.6		Control 2	<b>17.96</b>	±5.39
	Control 3	<b>46.2</b>	±13.9		Control 3	<b>39.83</b>	±11.95
DH-2000T series , DH-100T, DH-102T DH-103T, DH-105T DH-106T, DH-108T (Technical file version A2.0 or higher)	Control 1	<b>3.15</b>	±0.95	/	Control 1	/	/
	Control 2	<b>19.34</b>	±5.80		Control 2	/	/
	Control 3	<b>46.12</b>	±13.84		Control 3	/	/