



DM-RET

HEMATOLOGY CONTROL

Reference Values provided by DYMIND
CONTROL
 2025-03-13

 2025-06-10

Applicable Instruments	Parameter	Unit	LOT RET0425L	LOT RET0425N	LOT RET0425H
DYMIND DH-800 series DH-800 CRP series DH-800 CS series (Technical File Version A1.3 or higher)	RET#	$\times 10^{12}/L$	0.0374 ± 0.0187	0.1271 ± 0.0763	0.2345 ± 0.1173
	RET%	%	1.12 ± 0.56	3.83 ± 1.53	7.05 ± 2.82
	LFR	%	89.4 ± 27.0	89.4 ± 27.0	89.5 ± 27.0
	MFR	%	7.6 ± 11.0	7.5 ± 11.0	7.5 ± 11.0
	HFR	%	3.0 ± 5.0	3.1 ± 5.0	3.0 ± 5.0
	IRF	%	10.6 ± 16.0	10.6 ± 16.0	10.5 ± 16.0
	RHE	pg	29.1 ± 10.0	28.8 ± 10.0	28.8 ± 10.0

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0425L



RET0425N



RET0425H