	文件类型	技术文件	文件版本	V11.5.0	
	项目代号	/	页码	第0页 共2页	
	文件编号	GL DM-56-05-en			
文件名称： <p style="text-align: center;">CALIBRATOR N0525 2025-06-10</p> 适用范围： 公司所有三分类仪器					
编制：郭罕星 日期：2025.04.17		审核：刘金盈 日期：2025.04.17		批准：程军 日期：2025.04.17	
修 订 记 录					
版本	TCN/ECR/PCN编号	修订内容概要	修改人	批准人	批准日期
V11.5.0	/	新建	郭罕星	程军	2025.04.17
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由工程					


DM-CAL


HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT	N0525
------------	-------

	2025-04-12
---	------------

	2025-06-10
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31 Vet, DH36 Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	9.58
	RBC	$\times 10^{12}/L$	4.41
	HGB	g/L	131
	MCV	fL	94.1
	PLT	$\times 10^9/L$	238
DH21, DH23 DH25, DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.34
	RBC	$\times 10^{12}/L$	4.38
	HGB	g/L	129
	MCV	fL	85.8
	PLT	$\times 10^9/L$	236
DH20, DH22 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.70
	RBC	$\times 10^{12}/L$	4.49
	HGB	g/L	131
	MCV	fL	90.9
	PLT	$\times 10^9/L$	243
DH31 Vet, DH36 Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	9.31
	RBC	$\times 10^{12}/L$	4.35
	HGB	g/L	126
	MCV	fL	93.5
	PLT	$\times 10^9/L$	227

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.


DM-CAL


HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT	N0525
-----	-------

	2025-04-12
---	------------

	2025-06-10
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.51
	RBC	$\times 10^{12}/L$	4.39
	HGB	g/L	127
	MCV	fL	89.5
	PLT	$\times 10^9/L$	234
DP-H10 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.31
	RBC	$\times 10^{12}/L$	4.35
	HGB	g/L	136
	MCV	fL	93.5
	PLT	$\times 10^9/L$	227
DP-H10, DP-H12 DP-H13, DP-H15 DP-H16, DP-H17 (Technical File Version A3.3 or higher)	WBC	$\times 10^9/L$	10.03
	RBC	$\times 10^{12}/L$	4.61
	HGB	g/L	139
	MCV	fL	92.0
	PLT	$\times 10^9/L$	244

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.