

	文件类型	技术文件	文件版本	V3.5.0		
	项目代号	/	页码	第0页, 共1页		
	文件编号	GL DM-56-61-en				
文件名称: <p style="text-align: center;">CALIBRATOR CPF0525 2025-06-10</p> 适用范围: 1117项目						
编制: 郭罕星 日期: 2025.04.28		审核: 刘金盈 日期: 2025.04.28		批准: 程军 日期: 2025.04.28		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V3.5.0	/	新建		郭罕星	程军	2025.04.28
发至:	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由工程						

DM-CAL-PF

HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT CPF0525

2025-04-12

2025-06-10

Applicable Instruments	Parameter	Unit	Reference Value
DYMIND DH-800 series DH-800 CRP series DH-800 CS series (Technical File Version A1.3 or higher)	PLT-F	$\times 10^9/L$	202

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.