
	文件类型	技术文件	文件版本	V6.1.0		
	项目代号	/	页码	第0页 共2页		
	文件编号	GL DM-56-06-en				
文件名称： <p style="text-align: center;">Table of C-Reactive Protein Control (CRP) Reference Value</p> 适用范围： DM3113项目、DM3107项目、DM3513项目						
编制：郭罕星		审核：刘金盈		批准：程军		
日期：2025.04.30		日期：2025.04.30		日期：2025.04.30		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V6.1.0	/	新建		郭罕星	程军	2025.04.30
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由工程						

Table of C-Reactive Protein Control (CRP) Reference Value

C-Reactive Protein Control

LOT 2025041701


 2026-04-16

Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)	Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)
DH51CRP, DH53CRP, DH56CRP, D1-CRP, D3-CRP, D5-CRP (Technical file version A11.10 or higher with heating)	Control 1	2.9	±0.9	DH51CRP, DH53CRP, DH56CRP, D1-CRP, D3-CRP, D5-CRP (Technical file version A11.10 or higher without heating)	Control 1	3.0	±0.9
	Control 2	14.1	±4.2		Control 2	15.7	±4.7
	Control 3	33.1	±9.9		Control 3	38.3	±11.5
DH71CRP, DH73CRP, D2-CRP, D7-CRP (Technical file version A6.0 or higher with heating)	Control 1	3.1	±0.9	DH71CRP, DH73CRP, D2-CRP, D7-CRP (Technical file version A6.0 or higher without heating)	Control 1	2.7	±0.8
	Control 2	15.7	±4.7		Control 2	15.0	±4.5
	Control 3	41.2	±12.3		Control 3	35.3	±10.6
D6-CRP, DH76CRP (Technical file version B2.0)	Control 1	3.2	±1.0	D6-CRP, DH76CRP (Technical file version B2.2 or higher)	Control 1	2.8	±0.8
	Control 2	16.3	±4.9		Control 2	13.8	±4.1
	Control 3	41.1	±12.3		Control 3	34.6	±10.4
DF1-CRP, DF3-CRP, DF5-CRP, DF50CRP, DF52CRP, DF53CRP (Technical file version A6.0 or higher with heating)	Control 1	2.7	±0.8	DF1-CRP, DF3-CRP, DF5-CRP, DF50CRP, DF52CRP, DF53CRP (Technical file versions A8.1 and below without heating)	Control 1	3.3	±1.0
	Control 2	13.4	±4.0		Control 2	14.8	±4.4
	Control 3	32.7	±9.8		Control 3	35.8	±10.7

Table of C-Reactive Protein Control (CRP) Reference Value

C-Reactive Protein Control

LOT 2025041701

 2026-04-16

Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)	Applicable Instruments	Level	Reference Value (mg/L)	Deviation Range (mg/L)
DM79X, DM78X, DM72X (Technical file version A1.0 or higher)	Control 1	3.3	±1.0	DH36X, DH33X, DH31X (Technical file versions A1.0 or higher)	Control 1	2.9	±0.9
	Control 2	16.0	±4.8		Control 2	14.2	±4.3
	Control 3	37.3	±11.2		Control 3	34.2	±10.3
DP-H10, DP-H12, DP-H13, DP-H15, DP-H16, DP-H17 (Technical file versions A1.0 or higher)	Control 1	4.2	±1.3	DH-800 CRP series DH-800 CS series (Technical file Version A1.3 or higher)	Control 1	3.38	±1.01
	Control 2	18.8	±5.6		Control 2	16.27	±4.88
	Control 3	40.4	±12.1		Control 3	38.20	±11.46
DH-2000T series , DH-100T, DH-102T DH-103T, DH-105T DH-106T, DH-108T (Technical file version A2.0 or higher)	Control 1	3.25	±0.98	/	Control 1	/	/
	Control 2	15.69	±4.71		Control 2	/	/
	Control 3	36.62	±10.99		Control 3	/	/