



	文件类型	技术文件	文件版本	V5.3.0	
	项目代号	/	页码	第0页 共2页	
	文件编号	GL DM-56-18-en			
文件名称： <p style="text-align: center;">CONTROL DM-6D 6D0625 2025-08-10</p> 适用范围： 1116项目、1117项目					
编制：郭罕星 日期：2025.05.21		审核：刘金盈 日期：2025.05.21		批准：程军 日期：2025.05.21	
修 订 记 录					
版本	TCN/ECR/PCN编号	修订内容概要	修改人	批准人	批准日期
V5.3.0	/	新建	郭罕星	程军	2025.05.21
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检				
由工程					

DM-6D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND


CONTROL
 2025-05-13

 2025-08-10

Applicable Instruments	Parameter	Unit	LOT 6D0625L	LOT 6D0625N	LOT 6D0625H
DYMIND DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	WBC	$\times 10^9/L$	4.25 ± 1.20	7.90 ± 1.80	22.66 ± 3.00
	Neu#	$\times 10^9/L$	3.11 ± 0.40	5.19 ± 0.80	16.25 ± 2.00
	Lym#	$\times 10^9/L$	0.48 ± 0.20	1.79 ± 0.50	4.46 ± 1.40
	Mon#	$\times 10^9/L$	0.43 ± 0.20	0.67 ± 0.40	0.95 ± 0.80
	Eos#	$\times 10^9/L$	0.23 ± 0.20	0.25 ± 0.25	1.00 ± 0.90
	Bas#	$\times 10^9/L$	3.55 ± 0.40	6.08 ± 0.70	17.02 ± 1.60
	IG#	$\times 10^9/L$	0.58 ± 0.20	1.02 ± 0.30	3.08 ± 0.80
	Neu%	%	73.1 ± 8.0	65.7 ± 7.0	71.7 ± 8.0
	Lym%	%	11.4 ± 5.0	22.6 ± 7.0	19.7 ± 6.0
	Mon%	%	10.1 ± 6.0	8.5 ± 5.0	4.2 ± 4.0
	Eos%	%	5.4 ± 4.0	3.2 ± 3.2	4.4 ± 4.0
	Bas%	%	83.5 ± 10.0	77.0 ± 9.0	75.1 ± 9.0
	IG%	%	13.7 ± 4.5	12.9 ± 4.7	13.6 ± 4.5
	RBC	$\times 10^{12}/L$	2.25 ± 0.20	4.43 ± 0.30	5.31 ± 0.50
	HGB	g/L	63 ± 4	132 ± 6	173 ± 8
	HCT	%	17.6 ± 2.0	37.6 ± 2.5	49.0 ± 3.0
	MCV	fL	78.3 ± 5.0	84.9 ± 5.0	92.3 ± 5.0
	MCH	pg	28.0 ± 2.5	29.8 ± 2.5	32.6 ± 2.5
	MCHC	g/L	358 ± 30	351 ± 30	353 ± 30
	RDW-CV	%	17.8 ± 5.0	15.7 ± 5.0	15.2 ± 6.0
	RDW-SD	fL	50.7 ± 10.0	49.1 ± 10.0	51.0 ± 12.0
	PLT	$\times 10^9/L$	70 ± 20	292 ± 45	544 ± 65
	MPV	fL	7.2 ± 3.0	7.2 ± 3.0	7.4 ± 3.0
	PDW	fL	8.5 ± 5.0	8.5 ± 5.0	8.6 ± 5.0
	PCT	%	0.050 ± 0.050	0.210 ± 0.100	0.403 ± 0.200
	P-LCR	%	10.9 ± 10.9	10.0 ± 10.0	11.3 ± 11.3
	P-LCC	$\times 10^9/L$	8 ± 8	29 ± 29	61 ± 50
	IPF	%	2.2 ± 2.2	1.9 ± 1.9	1.7 ± 1.7

【NOTE】


1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



6D0625L



6D0625N





6D0625H

DM-6D

HEMATOLOGY CONTROL

Reference Values provided by DYMIND


CONTROL
 2025-05-13

 2025-08-10


Applicable Instruments	Parameter	Unit	LOT 6D0625L	LOT 6D0625N	LOT 6D0625H
DYMIND DH-800 series DH-800 CRP series DH-800 CS series (Technical File Version A1.3 or higher)	WBC	$\times 10^9/L$	3.90 ± 1.20	7.46 ± 1.80	20.84 ± 3.00
	Neu#	$\times 10^9/L$	2.83 ± 0.40	4.84 ± 0.80	14.95 ± 2.00
	Lym#	$\times 10^9/L$	0.41 ± 0.20	1.65 ± 0.50	3.81 ± 1.40
	Mon#	$\times 10^9/L$	0.37 ± 0.20	0.58 ± 0.40	0.83 ± 0.80
	Eos#	$\times 10^9/L$	0.19 ± 0.19	0.20 ± 0.20	0.79 ± 0.79
	Bas#	$\times 10^9/L$	0.10 ± 0.10	0.19 ± 0.19	0.46 ± 0.46
	IG#	$\times 10^9/L$	0.55 ± 0.20	0.98 ± 0.30	2.94 ± 0.80
	Neu%	%	72.7 ± 8.0	64.8 ± 7.0	71.7 ± 8.0
	Lym%	%	10.5 ± 5.0	22.1 ± 7.0	18.3 ± 6.0
	Mon%	%	9.4 ± 6.0	7.8 ± 5.0	4.0 ± 4.0
	Eos%	%	4.9 ± 4.0	2.7 ± 2.7	3.8 ± 3.8
	Bas%	%	2.5 ± 2.5	2.6 ± 2.6	2.2 ± 2.2
	IG%	%	14.2 ± 4.5	13.2 ± 4.7	14.1 ± 4.5
	RBC	$\times 10^{12}/L$	2.35 ± 0.20	4.57 ± 0.30	5.47 ± 0.50
	HGB	g/L	61 ± 4	132 ± 6	173 ± 8
	HCT	%	19.3 ± 2.0	40.5 ± 2.5	52.9 ± 3.0
	MCV	fL	82.2 ± 5.0	88.7 ± 5.0	96.6 ± 5.0
	MCH	pg	26.0 ± 2.5	28.9 ± 2.5	31.6 ± 2.5
	MCHC	g/L	316 ± 30	326 ± 30	327 ± 30
	RDW-CV	%	20.0 ± 5.0	16.9 ± 5.0	16.2 ± 6.0
	RDW-SD	fL	54.6 ± 10.0	51.1 ± 10.0	52.9 ± 12.0
	PLT	$\times 10^9/L$	74 ± 20	279 ± 45	511 ± 65
	MPV	fL	8.2 ± 3.0	8.0 ± 3.0	8.3 ± 3.0
	PDW	fL	7.9 ± 5.0	7.8 ± 5.0	7.8 ± 5.0
	PCT	%	0.061 ± 0.050	0.223 ± 0.100	0.424 ± 0.200
	P-LCR	%	13.2 ± 13.2	12.0 ± 12.0	13.1 ± 13.1
	P-LCC	$\times 10^9/L$	10 ± 10	34 ± 30	67 ± 50
	IPF	%	12.5 ± 5.0	12.5 ± 5.0	12.4 ± 5.0
NRBC#	$\times 10^9/L$	0.20 ± 0.20	0.02 ± 0.02	1.40 ± 0.40	
NRBC%	%	5.0 ± 4.0	0.3 ± 0.3	6.7 ± 4.0	

【NOTE】


- The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
- Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
- Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
- After using, put the controls back into the refrigerator to prevent contamination and evaporation.



6D0625L



6D0625N



6D0625H