

	文件类型	技术文件	文件版本	V5.7.0		
	项目代号	/	页码	第0页 共2页		
	文件编号	GL DM-56-17-en				
文件名称： <p style="text-align: center;">CONTROL DM-RET RET0725 2025-09-10</p> 适用范围： 1116项目、7003项目						
编制：刘金盈 日期：2025.07.01		审核：程军 日期：2025.07.01		批准：程军 日期：2025.07.01		
修 订 记 录						
版本	TCN/ECR/PCN编号	修订内容概要		修改人	批准人	批准日期
V5.7.0	/	新建		刘金盈	程军	2025.07.01
发至：	<input type="checkbox"/> 总经办 <input type="checkbox"/> 质量部 <input type="checkbox"/> 财务部 <input checked="" type="checkbox"/> 营销部 <input checked="" type="checkbox"/> 客服 <input type="checkbox"/> 计划 <input type="checkbox"/> 采购 <input checked="" type="checkbox"/> 工程 <input type="checkbox"/> 仓库 <input checked="" type="checkbox"/> 设备研发 <input checked="" type="checkbox"/> 试剂研发 <input type="checkbox"/> 物料质量 <input type="checkbox"/> 设备生产 <input type="checkbox"/> 试剂生产 <input checked="" type="checkbox"/> 试剂质检 <input checked="" type="checkbox"/> 设备质检					
由工程						

DM-RET

HEMATOLOGY CONTROL

Reference Values provided by DYMIND
CONTROL
 2025-06-13

 2025-09-10

Applicable Instruments	Parameter	Unit	LOT RET0725L	LOT RET0725N	LOT RET0725H
DYMIND DH-600 DH-602 DH-605 DH-610 DH-612 DH-615 (Technical File Version A2.0 or higher)	RET#	$\times 10^{12}/L$	0.0352 \pm 0.0176	0.2181 \pm 0.1309	0.2112 \pm 0.1056
	RET%	%	0.65 \pm 0.33	4.84 \pm 1.94	9.19 \pm 3.68
	LFR	%	90.0 \pm 27.0	90.3 \pm 27.0	89.5 \pm 27.0
	MFR	%	7.8 \pm 12.0	7.7 \pm 12.0	8.5 \pm 13.0
	HFR	%	2.2 \pm 3.0	2.0 \pm 3.0	2.0 \pm 3.0
	IRF	%	10.0 \pm 15.0	9.7 \pm 15.0	10.5 \pm 16.0
	RHE	pg	18.9 \pm 8.0	16.7 \pm 7.0	16.3 \pm 7.0

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0725L




RET0725N




RET0725H

DM-RET

HEMATOLOGY CONTROL

Reference Values provided by DYMIND
CONTROL
 2025-06-13

 2025-09-10

Applicable Instruments	Parameter	Unit	LOT RET0725L	LOT RET0725N	LOT RET0725H
DYMIND DM60 Vet DM61 Vet DM62 Vet DM63 Vet (Technical File Version A1.4 or higher)	RET#	$\times 10^{12}/L$	0.0099 ±0.0079	0.1387 ±0.0694	0.1053 ±0.0527
	RET%	%	0.19 ±0.15	3.20 ±1.60	4.68 ±2.34

【NOTE】

1. The controls should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 14 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the control after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Controls must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the controls back into the refrigerator to prevent contamination and evaporation.



RET0725L



RET0725N



RET0725H