



Coagulation Control Reference Value Table

Coagulation Control-level 1

CONTROL

LOT	2025021401
-----	------------

	20250214
---	----------

	20270213
---	----------

Applicable Instrument	Parameter	Unit	Principle	Reference Value	Deviation Limit
CA1200 CA1203 CA1205 DC1210 DC1250 DC1280	PT	s	Optical method	11.4	± 2.3
		s	Magnetic bead	11.6	± 2.3
	APTT	s	Optical method	31.0	± 6.2
		s	Magnetic bead	28.2	± 5.6
	TT	s	Optical method	15.6	± 3.1
		s	Magnetic bead	17.6	± 3.5
	FIB	g/L	Optical method	3.39	± 0.68
		g/L	Magnetic bead	3.23	± 0.65
AT-III	%	Optical method	83.8	± 16.8	

【NOTE】

1. The calibrator should be stored in the refrigerator at 2 °C - 8 °C. After reconstitution, the calibrator can be stable for 4 hours at room temperature (10 °C - 30 °C), and can be stable for 24 hours at 2 °C - 8 °C.
2. The calibrator cannot be frozen after reconstitution
3. The calibrator is for IVD only.